

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: ENTACT **SDG No.:** Q1641
Contract: ENTA05 **Lab Code:** CHEM **Case No.:** Q1641 **SAS No.:** Q1641

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB167355BS							
Aluminum	ug/L	1000	905		90	80 - 120	P
Arsenic	ug/L	400	370		92	80 - 120	P
Barium	ug/L	100	83.4		83	80 - 120	P
Cadmium	ug/L	100	91.2		91	80 - 120	P
Chromium	ug/L	200	195		98	80 - 120	P
Cobalt	ug/L	100	92.2		92	80 - 120	P
Copper	ug/L	150	148		99	80 - 120	P
Iron	ug/L	1500	1370		91	80 - 120	P
Lead	ug/L	500	453		91	80 - 120	P
Nickel	ug/L	250	234		94	80 - 120	P
Selenium	ug/L	1000	941		94	80 - 120	P
Silver	ug/L	37.5	36.1		96	80 - 120	P
Tin	ug/L	350	325		93	80 - 120	P
Zinc	ug/L	100	98.5		98	80 - 120	P

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: ENTACT

SDG No.: Q1641

Contract: ENTA05

Lab Code: CHEM

Case No.: Q1641

SAS No.: Q1641

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB167356BS Mercury	ug/L	4.0	4.33		108	82 - 119	CV