

CALIBRATION VERIFICATION SUMMARY

Contract: ENTA05

Lab Code: CHEM Case No.: Q1641 SAS No.: Q1641 SDG NO.: Q1641

Continuing Calib Date: 03/26/2025 Initial Calibration Date(s): 03/11/2025 03/11/2025

Continuing Calib Time: 17:03 Initial Calibration Time(s): 15:13 23:38

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Aroclor-1248-1 (1)	5.67	5.68	5.58	5.78	0.01
Aroclor-1248-2 (2)	5.94	5.95	5.85	6.05	0.01
Aroclor-1248-3 (3)	6.15	6.15	6.05	6.25	0.01
Aroclor-1248-4 (4)	6.54	6.55	6.45	6.65	0.01
Aroclor-1248-5 (5)	6.58	6.59	6.49	6.69	0.01
Tetrachloro-m-xylene	4.52	4.52	4.42	4.62	0.00
Decachlorobiphenyl	10.24	10.25	10.15	10.35	0.01

CALIBRATION VERIFICATION SUMMARY

Contract: ENTA05

Lab Code: CHEM Case No.: Q1641 SAS No.: Q1641 SDG NO.: Q1641

Continuing Calib Date: 03/26/2025 Initial Calibration Date(s): 03/11/2025 03/11/2025

Continuing Calib Time: 17:03 Initial Calibration Time(s): 15:13 23:38

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Aroclor-1248-1 (1)	4.91	4.92	4.82	5.02	0.01
Aroclor-1248-2 (2)	5.14	5.15	5.05	5.25	0.01
Aroclor-1248-3 (3)	5.19	5.20	5.10	5.30	0.02
Aroclor-1248-4 (4)	5.36	5.37	5.27	5.47	0.01
Aroclor-1248-5 (5)	5.75	5.76	5.66	5.86	0.01
Tetrachloro-m-xylene	3.82	3.83	3.73	3.93	0.01
Decachlorobiphenyl	8.86	8.88	8.78	8.98	0.02