

CALIBRATION VERIFICATION SUMMARY

Contract: ENTA05

Lab Code: CHEM Case No.: Q1641 SAS No.: Q1641 SDG NO.: Q1641

Continuing Calib Date: 03/26/2025 Initial Calibration Date(s): 03/11/2025 03/11/2025

Continuing Calib Time: 21:33 Initial Calibration Time(s): 15:13 23:38

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW FROM TO		DIFF RT
Aroclor-1016-1	(1)	5.67	5.68	5.58	5.78	0.01
Aroclor-1016-2	(2)	5.69	5.70	5.60	5.80	0.01
Aroclor-1016-3	(3)	5.75	5.76	5.66	5.86	0.01
Aroclor-1016-4	(4)	5.85	5.86	5.76	5.96	0.01
Aroclor-1016-5	(5)	6.14	6.15	6.05	6.25	0.01
Aroclor-1260-1	(1)	7.26	7.27	7.17	7.37	0.01
Aroclor-1260-2	(2)	7.51	7.53	7.43	7.63	0.02
Aroclor-1260-3	(3)	7.87	7.88	7.78	7.98	0.01
Aroclor-1260-4	(4)	8.10	8.11	8.01	8.21	0.01
Aroclor-1260-5	(5)	8.42	8.43	8.33	8.53	0.01
Tetrachloro-m-xylene		4.52	4.52	4.42	4.62	0.00
Decachlorobiphenyl		10.23	10.25	10.15	10.35	0.02

CALIBRATION VERIFICATION SUMMARY

Contract: ENTA05

Lab Code: CHEM Case No.: Q1641 SAS No.: Q1641 SDG NO.: Q1641

Continuing Calib Date: 03/26/2025 Initial Calibration Date(s): 03/11/2025 03/11/2025

Continuing Calib Time: 21:33 Initial Calibration Time(s): 15:13 23:38

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1016-1	(1)	4.91	4.92	4.82	5.02	0.01
Aroclor-1016-2	(2)	4.92	4.94	4.84	5.04	0.02
Aroclor-1016-3	(3)	5.10	5.11	5.01	5.21	0.01
Aroclor-1016-4	(4)	5.14	5.15	5.05	5.25	0.01
Aroclor-1016-5	(5)	5.36	5.37	5.27	5.47	0.01
Aroclor-1260-1	(1)	6.39	6.41	6.31	6.51	0.02
Aroclor-1260-2	(2)	6.58	6.60	6.50	6.70	0.02
Aroclor-1260-3	(3)	6.73	6.75	6.65	6.85	0.02
Aroclor-1260-4	(4)	7.21	7.22	7.12	7.32	0.01
Aroclor-1260-5	(5)	7.45	7.46	7.36	7.56	0.01
Tetrachloro-m-xylene		3.82	3.83	3.73	3.93	0.01
Decachlorobiphenyl		8.86	8.88	8.78	8.98	0.02