

CALIBRATION VERIFICATION SUMMARY

Contract: ENTA05

Lab Code: CHEM Case No.: Q1641 SAS No.: Q1641 SDG NO.: Q1641

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 03/11/2025 03/11/2025

Client Sample No.: CCAL01 Date Analyzed: 03/26/2025

Lab Sample No.: AR1660CCC500 Data File : PP070896.D Time Analyzed: 16:14

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	5.670	5.577	5.777	451.950	500.000	-9.6
Aroclor-1016-2	5.691	5.599	5.799	495.890	500.000	-0.8
Aroclor-1016-3	5.754	5.661	5.861	469.060	500.000	-6.2
Aroclor-1016-4	5.851	5.758	5.958	449.900	500.000	-10.0
Aroclor-1016-5	6.144	6.051	6.251	439.900	500.000	-12.0
Aroclor-1260-1	7.263	7.171	7.371	582.920	500.000	16.6
Aroclor-1260-2	7.516	7.425	7.625	583.730	500.000	16.7
Aroclor-1260-3	7.874	7.783	7.983	564.940	500.000	13.0
Aroclor-1260-4	8.098	8.008	8.208	515.310	500.000	3.1
Aroclor-1260-5	8.419	8.328	8.528	524.900	500.000	5.0
Decachlorobiphenyl	10.233	10.152	10.352	48.990	50.000	-2.0
Tetrachloro-m-xylene	4.518	4.424	4.624	47.670	50.000	-4.7

CALIBRATION VERIFICATION SUMMARY

Contract: ENTA05

Lab Code: CHEM Case No.: Q1641 SAS No.: Q1641 SDG NO.: Q1641

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 03/11/2025 03/11/2025

Client Sample No.: CCAL01 Date Analyzed: 03/26/2025

Lab Sample No.: AR1660CCC500 Data File : PP070896.D Time Analyzed: 16:14

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.905	4.817	5.017	460.620	500.000	-7.9
Aroclor-1016-2	4.924	4.836	5.036	461.270	500.000	-7.7
Aroclor-1016-3	5.101	5.013	5.213	471.720	500.000	-5.7
Aroclor-1016-4	5.143	5.054	5.254	489.260	500.000	-2.1
Aroclor-1016-5	5.358	5.270	5.470	494.350	500.000	-1.1
Aroclor-1260-1	6.394	6.307	6.507	471.420	500.000	-5.7
Aroclor-1260-2	6.582	6.495	6.695	456.520	500.000	-8.7
Aroclor-1260-3	6.735	6.649	6.849	454.060	500.000	-9.2
Aroclor-1260-4	7.206	7.121	7.321	434.050	500.000	-13.2
Aroclor-1260-5	7.448	7.362	7.562	429.530	500.000	-14.1
Decachlorobiphenyl	8.861	8.781	8.981	41.740	50.000	-16.5
Tetrachloro-m-xylene	3.820	3.729	3.929	49.800	50.000	-0.4