

CALIBRATION VERIFICATION SUMMARY

Contract: ENTA05

Lab Code: CHEM Case No.: Q1641 SAS No.: Q1641 SDG NO.: Q1641

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 03/11/2025 03/11/2025

Client Sample No.: CCAL02 Date Analyzed: 03/26/2025

Lab Sample No.: AR1660CCC500 Data File : PP070910.D Time Analyzed: 21:33

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	5.668	5.577	5.777	440.830	500.000	-11.8
Aroclor-1016-2	5.690	5.599	5.799	481.230	500.000	-3.8
Aroclor-1016-3	5.753	5.661	5.861	446.120	500.000	-10.8
Aroclor-1016-4	5.850	5.758	5.958	440.340	500.000	-11.9
Aroclor-1016-5	6.142	6.051	6.251	431.200	500.000	-13.8
Aroclor-1260-1	7.259	7.171	7.371	528.370	500.000	5.7
Aroclor-1260-2	7.512	7.425	7.625	566.600	500.000	13.3
Aroclor-1260-3	7.871	7.783	7.983	567.000	500.000	13.4
Aroclor-1260-4	8.095	8.008	8.208	537.430	500.000	7.5
Aroclor-1260-5	8.416	8.328	8.528	550.550	500.000	10.1
Decachlorobiphenyl	10.231	10.152	10.352	48.350	50.000	-3.3
Tetrachloro-m-xylene	4.517	4.424	4.624	46.810	50.000	-6.4

CALIBRATION VERIFICATION SUMMARY

Contract: ENTA05

Lab Code: CHEM Case No.: Q1641 SAS No.: Q1641 SDG NO.: Q1641

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 03/11/2025 03/11/2025

Client Sample No.: CCAL02 Date Analyzed: 03/26/2025

Lab Sample No.: AR1660CCC500 Data File : PP070910.D Time Analyzed: 21:33

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.905	4.817	5.017	426.080	500.000	-14.8
Aroclor-1016-2	4.924	4.836	5.036	459.950	500.000	-8.0
Aroclor-1016-3	5.101	5.013	5.213	454.940	500.000	-9.0
Aroclor-1016-4	5.143	5.054	5.254	456.480	500.000	-8.7
Aroclor-1016-5	5.357	5.270	5.470	444.590	500.000	-11.1
Aroclor-1260-1	6.393	6.307	6.507	460.920	500.000	-7.8
Aroclor-1260-2	6.581	6.495	6.695	449.170	500.000	-10.2
Aroclor-1260-3	6.734	6.649	6.849	439.200	500.000	-12.2
Aroclor-1260-4	7.206	7.121	7.321	434.850	500.000	-13.0
Aroclor-1260-5	7.447	7.362	7.562	436.900	500.000	-12.6
Decachlorobiphenyl	8.860	8.781	8.981	44.810	50.000	-10.4
Tetrachloro-m-xylene	3.820	3.729	3.929	47.450	50.000	-5.1