

Surrogate Summary

SDG No.: Q1641

Client: ENTACT

Analytical Method: SW8260D

Lab Sample ID	Client ID	Parameter	Spike	Result	RecoveryQual	Limits	
						Low	High
Q1641-02	NP-WS-003	1,2-Dichloroethane-d4	50	54.4	109	70 (74)	130 (125)
		Dibromofluoromethane	50	50.1	100	70 (75)	130 (124)
		Toluene-d8	50	47.4	95	70 (86)	130 (113)
		4-Bromofluorobenzene	50	45.5	91	70 (77)	130 (121)

Surrogate Summary

SDG No.: Q1641

Client: ENTACT

Analytical Method: SW8260-Low

Lab Sample ID	Client ID	Parameter	Spike	Result	RecoveryQual	Limits	
						Low	High
VN0326WBL01	VN0326WBL01	1,2-Dichloroethane-d4	50	58.2	116	70 (74)	130 (125)
		Dibromofluoromethane	50	53.8	108	70 (75)	130 (124)
		Toluene-d8	50	48.6	97	70 (86)	130 (113)
		4-Bromofluorobenzene	50	43.0	86	70 (77)	130 (121)
VN0326WBS01	VN0326WBS01	1,2-Dichloroethane-d4	50	50.2	100	70 (74)	130 (125)
		Dibromofluoromethane	50	52.2	104	70 (75)	130 (124)
		Toluene-d8	50	53.5	107	70 (86)	130 (113)
		4-Bromofluorobenzene	50	50.1	100	70 (77)	130 (121)
VN0326WBSD0	VN0326WBSD01	1,2-Dichloroethane-d4	50	49.1	98	70 (74)	130 (125)
		Dibromofluoromethane	50	49.6	99	70 (75)	130 (124)
		Toluene-d8	50	51.3	103	70 (86)	130 (113)
		4-Bromofluorobenzene	50	50.0	100	70 (77)	130 (121)

