

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: ENTA05
 Lab Code: CHEM Case No.: Q1641 SAS No.: Q1641 SDG NO.: Q1641
 Lab File ID: VN086118.D Date Analyzed: 03/26/2025
 Instrument ID: MSVOA_N Time Analyzed: 11:38
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	196436	8.22	311095	9.10	292741	11.87
UPPER LIMIT	392872	8.724	622190	9.6	585482	12.365
LOWER LIMIT	98218	7.724	155548	8.6	146371	11.365
EPA SAMPLE NO.						
NP-WS-003	168751	8.22	312504	9.10	275427	11.87
VN0326WBL01	165546	8.22	311785	9.10	280557	11.87
VN0326WBS01	204522	8.22	330210	9.10	294866	11.87
VN0326WBSD01	217618	8.22	354450	9.10	311562	11.87

IS1 = Pentafluorobenzene
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.