

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q1649

Client: GFE LLC

Analytical Method: SWTO-15

Datafile : VL042203.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VL0331ABS01	Dichlorodifluoromethane	10	10.4	ppbv	104			70	130	
	Chloromethane	10	9.70	ppbv	97			70	130	
	Vinyl Chloride	10	8.70	ppbv	87			70	130	
	Bromomethane	10	9.10	ppbv	91			70	130	
	Chloroethane	10	9.30	ppbv	93			70	130	
	Tetrahydrofuran	10	9.30	ppbv	93			70	130	
	Trichlorofluoromethane	10	10.3	ppbv	103			70	130	
	1,1,2-Trichlorotrifluoroethane	10	9.70	ppbv	97			70	130	
	Dichlorotetrafluoroethane	10	10.3	ppbv	103			70	130	
	tert-Butyl Alcohol	10	9.00	ppbv	90			70	130	
	Heptane	10	9.20	ppbv	92			70	130	
	1,1-Dichloroethene	10	9.20	ppbv	92			70	130	
	Acetone	10	8.90	ppbv	89			70	130	
	Carbon disulfide	10	9.30	ppbv	93			70	130	
	Methyl tert-butyl Ether	10	11.0	ppbv	110			70	130	
	Methylene Chloride	10	8.40	ppbv	84			70	130	
	trans-1,2-Dichloroethene	10	9.20	ppbv	92			70	130	
	1,1-Dichloroethane	10	10.0	ppbv	100			70	130	
	Cyclohexane	10	9.40	ppbv	94			70	130	
	2-Butanone	10	9.50	ppbv	95			70	130	
	Carbon Tetrachloride	10	10.6	ppbv	106			70	130	
	cis-1,2-Dichloroethene	10	10.0	ppbv	100			70	130	
	Chloroform	10	10.6	ppbv	106			70	130	
	1,1,1-Trichloroethane	10	9.80	ppbv	98			70	130	
	2,2,4-Trimethylpentane	10	9.50	ppbv	95			70	130	
	Benzene	10	9.80	ppbv	98			70	130	
	1,2-Dichloroethane	10	11.0	ppbv	110			70	130	
	Trichloroethene	10	9.70	ppbv	97			70	130	
	1,2-Dichloropropane	10	9.70	ppbv	97			70	130	
	Bromodichloromethane	10	10.8	ppbv	108			70	130	
	4-Methyl-2-Pentanone	10	8.60	ppbv	86			70	130	
	Toluene	10	9.60	ppbv	96			70	130	
	t-1,3-Dichloropropene	10	9.10	ppbv	91			70	130	
	cis-1,3-Dichloropropene	10	9.40	ppbv	94			70	130	
	1,1,2-Trichloroethane	10	10.2	ppbv	102			70	130	
	Dibromochloromethane	10	10.6	ppbv	106			70	130	
	1,2-Dibromoethane	10	10.3	ppbv	103			70	130	
	Tetrachloroethene	10	9.00	ppbv	90			70	130	
	Chlorobenzene	10	10.4	ppbv	104			70	130	
	Ethyl Benzene	10	10.1	ppbv	101			70	130	
	m/p-Xylene	20	20.6	ppbv	103			70	130	
	o-Xylene	10	10.4	ppbv	104			70	130	
	Styrene	10	10.1	ppbv	101			70	130	
	Bromoform	10	10.3	ppbv	103			70	130	
	1,1,2,2-Tetrachloroethane	10	10.4	ppbv	104			70	130	
	2-Chlorotoluene	10	10.3	ppbv	103			70	130	
	1,3,5-Trimethylbenzene	10	10.4	ppbv	104			70	130	
	1,2,4-Trimethylbenzene	10	9.80	ppbv	98			70	130	

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q1649

Client: GFE LLC

Analytical Method: SWTO-15

Datafile : VL042203.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VL0331ABS01	1,3-Dichlorobenzene	10	10.5	ppbv	105			70	130	
	1,4-Dichlorobenzene	10	10.5	ppbv	105			70	130	
	1,2-Dichlorobenzene	10	10.6	ppbv	106			70	130	
	1,2,4-Trichlorobenzene	10	9.30	ppbv	93			70	130	
	Hexachloro-1,3-butadiene	10	8.40	ppbv	84			70	130	
	Naphthalene	10	9.50	ppbv	95			70	130	
	1,3-Butadiene	10	8.70	ppbv	87			70	130	
	4-Ethyltoluene	10	10.5	ppbv	105			70	130	
	Hexane	10	9.50	ppbv	95			70	130	
	Allyl Chloride	10	8.90	ppbv	89			70	130	
	1,4-Dioxane	10	8.90	ppbv	89			70	130	
	Methyl methacrylate	10	9.20	ppbv	92			70	130	