

VOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
BROMOFLUOROBENZENE (BFB)

Lab Name: CHEMTECH Contract: GFEL01
Lab Code: CHEM Case No.: Q1649 SAS No.: Q1649 SDG NO.: Q1649
Lab File ID: VL042171.D BFB Injection Date: 03/27/2025
Instrument ID: MSVOA_L BFB Injection Time: 07:44
GC Column: RTX-1 ID: 0.32 (mm) Heated Purge: Y/N N

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
50	8.0 - 40.0% of mass 95	23
75	30.0 - 66.0% of mass 95	56.3
95	Base Peak, 100% relative abundance	100
96	5.0 - 9.0% of mass 95	6.6
173	Less than 2.0% of mass 174	0.0 (0.0) 1
174	50.0 - 120.0% of mass 95	63.8
175	4.0 - 9.0% of mass 174	5.1 (8) 1
176	93.0 - 101.0% of mass 174	61.1 (95.7) 1
177	5.0 - 9.0% of mass 176	4.2 (6.8) 2

1-Value is % mass 174

2-Value is % mass 176

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
VSTDICCC010	VSTDICCC010	VL042172.D	03/27/2025	08:26
VSTDICCC002	VSTDICCC002	VL042173.D	03/27/2025	08:59
VSTDICCC001	VSTDICCC001	VL042174.D	03/27/2025	09:30
VSTDICCC0.5	VSTDICCC0.5	VL042175.D	03/27/2025	10:01
VSTDICCC0.1	VSTDICCC0.1	VL042176.D	03/27/2025	10:32
VSTDICCC0.03	VSTDICCC0.03	VL042177.D	03/27/2025	11:03
VSTDICCC015	VSTDICCC015	VL042178.D	03/27/2025	11:36