284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	PSEG	Date Collected:	03/26/25
Project:	Livingston Pole Replacement	Date Received:	03/26/25
Client Sample ID:	STOCK-PILE-BIN2	SDG No.:	Q1650
Lab Sample ID:	Q1650-03	Matrix:	Solid
Analytical Method:	NJEPH	% Solid:	79.6
Sample Wt/Vol:	30.05 Units: g	Final Vol:	2000 uL
Soil Aliquot Vol:	uL	Test:	EPH_NF
Prep Method :			

Prep Date :	Date Analyzed :	Prep Batch ID
03/27/25 09:55	03/27/25 16:22	PB167350

CAS Number	Parameter	Conc.	Qualifier	Dilution	MDL	LOQ / CRQL	Units(Dry Weight)	Datafile
TARGETS								
Aliphatic C28-C40	Aliphatic C28-C40	6.92		1	1.48	2.51	mg/kg	FE053020.D
Aliphatic C9-C28	Aliphatic C9-C28	3.76	J	1	1.14	5.01	mg/kg	FE053020.D
Total AliphaticEPH	Total AliphaticEPH	10.7			2.62	7.52	mg/kg	
Total EPH	Total EPH	10.7			2.62	7.52	mg/kg	

* As samples are not fractionated, all aliphatic and aromatic carbon compounds in the C9-C40 carbon range are calculated against the aliphatic calibration curve, and reported as Aliphatic EPH. Therefore, the aliphatic C9-C40 concentration for the sample is reported as the Total EPH.

U = Not Detected
LOQ = Limit of Quantitation
MDL = Method Detection Limit
LOD = Limit of Detection
E = Value Exceeds Calibration Range
Q = indicates LCS control criteria did not meet requirements

J = Estimated Value
B = Analyte Found in Associated Method Blank
N = Presumptive Evidence of a Compound
* = Values outside of QC limits
D = Dilution

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	PSEG	Date Collected:	03/26/25
Project:	Livingston Pole Replacement	Date Received:	03/26/25
Client Sample ID:	STOCK-PILE-BIN2	SDG No.:	Q1650
Lab Sample ID:	Q1650-03	Matrix:	Solid
Analytical Method:	NJEPH	% Solid:	79.6
Sample Wt/Vol:	30.05 Units: g	Final Vol:	2000 uL
Soil Aliquot Vol:	uL	Test:	EPH_NF
Prep Method :			

File ID :	Dilution:	Prep Date :	Date Analyzed :	Prep Batch ID
FE053020.D	1	03/27/25	03/27/25	PB167350

CAS Number	Parameter		Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS							
Aliphatic C9-C28	Aliphatic C9-C28		3.76	J	1.14	5.01	mg/kg
Aliphatic C28-C40	Aliphatic C28-C40		6.92		1.48	2.51	mg/kg
SURROGATES							
3383-33-2	1-chlorooctadecane (SURR)		42.2		40 - 140	84%	SPK: 50
84-15-1	ortho-Terphenyl (SURR)		40.1		40 - 140	80%	SPK: 50

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

Quantitation Report For Aliphatic EPH Range.

Lab Sample ID:	Q1650-03	Acq On:	27 Mar 2025 16:22
Client Sample ID:	STOCK-PILE-BIN2	Operator:	YP\AJ
Data file:	FE053020.D	Misc:	
Instrument:	FID_E	ALS Vial:	16
Dilution Factor:	1	Sample Multiplier:	1.00

Compound	R.T.		Response	Conc	highest_standard	Units
Aliphatic C9-C12	3.143	6.785	1488542	9.545	300	ug/ml
Aliphatic C12-C16	6.786	10.236	1308912	8.606	200	ug/ml
Aliphatic C16-C21	10.237	13.612	1245401	8.498	300	ug/ml
Aliphatic C21-C28	13.613	17.282	2598428	18.241	400	ug/ml
Aliphatic C28-C40	17.283	22.191	10560611	82.77	600	ug/ml
Aliphatic EPH	3.143	22.191	17201894	127.66		ug/ml
ortho-Terphenyl (SURR)	11.901	11.901	6755458	40.07		ug/ml
1-chlorooctadecane (SURR)	13.344	13.344	5127267	42.23		ug/ml
Aliphatic C9-C28	3.143	17.282	6641283	44.89	1200	ug/ml