

Laboratory Control Sample Summary

Client: Tully Environmental, Inc

SDG No.: Q1651

Project: Transfer Station - SPDES 2025

Run No.: LB135205

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB135205BS							
TSS	mg/L	550	533		97	1	90-110	03/27/2025