

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Tully Environmental, Inc **SDG No.:** Q1651
Contract: TULL01 **Lab Code:** CHEM **Case No.:** Q1651 **SAS No.:** Q1651
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV01	Copper	523	510	102	95 - 105	P	03/27/2025	09:49	LB135215
	Iron	9820	10000	98	95 - 105	P	03/27/2025	09:49	LB135215
	Lead	977	1000	98	95 - 105	P	03/27/2025	09:49	LB135215

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Tully Environmental, Inc **SDG No.:** Q1651
Contract: TULL01 **Lab Code:** CHEM **Case No.:** Q1651 **SAS No.:** Q1651
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
LLICV01	Copper	22.0	20.0	110	80 - 120	P	03/27/2025	09:58	LB135215
	Iron	82.3	100	82	80 - 120	P	03/27/2025	09:58	LB135215
	Lead	10.9	12.0	91	80 - 120	P	03/27/2025	09:58	LB135215

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Tully Environmental, Inc **SDG No.:** Q1651
Contract: TULL01 **Lab Code:** CHEM **Case No.:** Q1651 **SAS No.:** Q1651
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV01	Copper	1270	1250	101	90 - 110	P	03/27/2025	10:28	LB135215
	Iron	5050	5000	101	90 - 110	P	03/27/2025	10:28	LB135215
	Lead	5020	5000	100	90 - 110	P	03/27/2025	10:28	LB135215
CCV02	Copper	1290	1250	103	90 - 110	P	03/27/2025	11:18	LB135215
	Iron	5140	5000	103	90 - 110	P	03/27/2025	11:18	LB135215
	Lead	5080	5000	102	90 - 110	P	03/27/2025	11:18	LB135215
CCV03	Copper	1260	1250	101	90 - 110	P	03/27/2025	12:10	LB135215
	Iron	4850	5000	97	90 - 110	P	03/27/2025	12:10	LB135215
	Lead	5030	5000	101	90 - 110	P	03/27/2025	12:10	LB135215
CCV04	Copper	1250	1250	100	90 - 110	P	03/27/2025	13:15	LB135215
	Iron	5010	5000	100	90 - 110	P	03/27/2025	13:15	LB135215
	Lead	4910	5000	98	90 - 110	P	03/27/2025	13:15	LB135215
CCV05	Copper	1240	1250	100	90 - 110	P	03/27/2025	13:52	LB135215
	Iron	4840	5000	97	90 - 110	P	03/27/2025	13:52	LB135215
	Lead	4910	5000	98	90 - 110	P	03/27/2025	13:52	LB135215