



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 2b -

CRDL STANDARD FOR AA & ICP

Client: Tully Environmental, Inc **SDG No.:** Q1651
Contract: TULL01 **Lab Code:** CHEM **Case No.:** Q1651 **SAS No.:** Q1651
Initial Calibration Source: _____
Continuing Calibration Source: _____

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CRI01	Copper	22.0	20.0	110	40 - 160	P	03/27/2025	10:07	LB135215
	Iron	85.7	100	86	40 - 160	P	03/27/2025	10:07	LB135215
	Lead	11.0	12.0	92	40 - 160	P	03/27/2025	10:07	LB135215