

Metals

- 4 -

INTERFERENCE CHECK SAMPLE

Client: Tully Environmental, Inc **SDG No.:** Q1651
Contract: TULL01 **Lab Code:** CHEM **Case No.:** Q1651 **SAS No.:** Q1651
ICS Source: EPA **Instrument ID:** P4

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSA01	Copper	2.70	2.0	135	-18	22	03/27/2025	10:11	LB135215
	Iron	97000	101000	96	85600	116500	03/27/2025	10:11	LB135215
	Lead	4.40			-12	12	03/27/2025	10:11	LB135215
ICSAB01	Copper	506	511	99	434	588	03/27/2025	10:15	LB135215
	Iron	99400	99300	100	84400	114500	03/27/2025	10:15	LB135215
	Lead	53.4	49.0	109	37	61	03/27/2025	10:15	LB135215
ICSA	Copper	-8.00	2.0	400	-18	22	03/27/2025	10:19	LB135215
	Iron	97900	101000	97	85600	116500	03/27/2025	10:19	LB135215
	Lead	-28.2			-12	12	03/27/2025	10:19	LB135215
ICSAB	Copper	521	511	102	434	588	03/27/2025	10:24	LB135215
	Iron	96800	99300	98	84400	114500	03/27/2025	10:24	LB135215
	Lead	45.3	49.0	92	37	61	03/27/2025	10:24	LB135215