

metals
- 5a -
MATRIX SPIKE SUMMARY

client: Tully Environmental, Inc **level:** low **sdg no.:** Q1651
contract: TULL01 **lab code:** CHEM **case no.:** Q1651 **sas no.:** Q1651
matrix: Water **sample id:** Q1651-01 **client id:** 001-WILLETS-PT-BLVD(MAR)MS
Percent Solids for Sample: NA **Spiked ID:** Q1651-01MS **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit %R	Spiked Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Copper	ug/L	75 - 125	160		11.0		150	99		P
Iron	ug/L	75 - 125	6460		4840		1500	108		P
Lead	ug/L	75 - 125	475		9.12		500	93		P

metals
- 5a -
MATRIX SPIKE DUPLICATE SUMMARY

client: <u>Tully Environmental, Inc</u>	level: <u>low</u>	sdg no.: <u>Q1651</u>
contract: <u>TULL01</u>	lab code: <u>CHEM</u>	case no.: <u>Q1651</u> sas no.: <u>Q1651</u>
matrix: <u>Water</u>	sample id: <u>Q1651-01</u>	client id: <u>001-WILLETS-PT-BLVD(MAR)MSD</u>
Percent Solids for Sample: NA	Spiked ID: Q1651-01MSD	Percent Solids for Spike Sample: NA

Analyte	Units	Acceptance Limit %R	MSD Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Copper	ug/L	75 - 125	162		11.0		150	101		P
Iron	ug/L	75 - 125	6190		4840		1500	90		P
Lead	ug/L	75 - 125	477		9.12		500	94		P