

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: Tully Environmental, Inc

SDG No.: Q1651

Contract: TULL01

Lab Code: CHEM

Case No.: Q1651

SAS No.: Q1651

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB167340BS							
Copper	ug/L	150	149		99	85 - 115	P
Iron	ug/L	1500	1440		96	85 - 115	P
Lead	ug/L	500	458		92	85 - 115	P