



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

## **CASE NARRATIVE**

**Weston Solutions, Inc.**

**Project Name: RFP 905**

**Project # N/A**

**Chemtech Project # Q1664**

**Test Name: SPLP Mercury, SPLP ICP Metals**

### **A. Number of Samples and Date of Receipt:**

22 Solid samples were received on 03/27/2025.

### **B. Parameters:**

According to the Chain of Custody document, the following analyses were requested: Cyanide, EPH, Mercury, Metals ICP-TAL, METALS TAL+CN, PCB, Pesticide-TCL, SPLP BNA, SPLP Cyanide, SPLP Extraction, SPLP ICP Metals, SPLP Mercury, SPLP PCB, SPLP Pesticide, SPLP VOA, SPLP ZHE Ext, SVOC-TCL BNA -20 and VOC-TCLVOA-10. This data package contains results for SPLP Mercury, SPLP ICP Metals.

### **C. Analytical Techniques:**

The analysis of SPLP ICP Metals was based on method 6010D, digestion based on method 3050 (soils). The analysis of SPLP Mercury was based on method 7470A and digestion was based on method 7471B (soils).

### **D. QA/ QC Samples:**

The Holding Times were met for all analysis.

The Blank Spike met requirements for all samples.

The Duplicate analysis met criteria for all samples.

The Matrix Spike (P001-BBDGA-001-01MS) analysis met criteria for all samples except for Copper due to matrix interference.

The Matrix Spike Duplicate (P001-BBDGA-001-01MSD) analysis met criteria for all samples except for Copper due to matrix interference.

The Blank analysis did not indicate the presence of lab contamination.

The Calibration met the requirements.

The Serial Dilution met the acceptable requirements.

### **E. Additional Comments:**

#### **Calculation for ICP-AES Water Sample:**

$$\text{Concentration or Result } (\mu\text{g/L}) = C \times \frac{V_f}{V_i} \times \text{DF} \times 1000$$



Where,

C = Instrument value in ppm (The average of all replicate exposures)

Vf = Final digestion volume (mL)

Vi = Initial aliquot amount (mL) (Sample amount taken in prep)

DF = Dilution Factor

**Calculation for Hg Water Sample:**

Concentration or Result ( $\mu\text{g/L}$ ) = C x DF

Where,

C = Instrument response in  $\mu\text{g/L}$  from the calibration curve.

DF = Dilution Factor

**F. Additional Comments:**

---

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed

above. The laboratory manager or his designee, as verified by the following signature has authorized release of the data contained in this hard copy data package.

Signature\_\_\_\_\_