

Initial and Continuing Calibration Verification

Client: Weston Solutions, Inc.

SDG No.: Q1664

Project: RFP 905

RunNo.: LB135233

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV1 Cyanide	mg/L	0.093	0.099	94	90-110	03/31/2025
Sample ID: CCV1 Cyanide	mg/L	0.24	0.25	96	90-110	03/31/2025
Sample ID: CCV2 Cyanide	mg/L	0.24	0.25	96	90-110	03/31/2025
Sample ID: CCV3 Cyanide	mg/L	0.25	0.25	100	90-110	03/31/2025

Initial and Continuing Calibration Verification

Client: Weston Solutions, Inc.

SDG No.: Q1664

Project: RFP 905

RunNo.: LB135243

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV1 Cyanide	mg/L	0.095	0.099	96	90-110	03/31/2025
Sample ID: CCV1 Cyanide	mg/L	0.24	0.25	96	90-110	03/31/2025
Sample ID: CCV2 Cyanide	mg/L	0.24	0.25	96	90-110	03/31/2025
Sample ID: CCV3 Cyanide	mg/L	0.25	0.25	100	90-110	03/31/2025