



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Initial and Continuing Calibration Blank Summary

Client: Weston Solutions, Inc.

SDG No.: Q1664

Project: RFP 905

RunNo.: LB135233

Analyte	Units	Result	Acceptance Limits	Conc Qual	MDL	RDL	Analysis Date
Sample ID: ICB1 Cyanide	mg/L	0.0017	0.0025	J	0.00096	0.005	03/31/2025
Sample ID: CCB1 Cyanide	mg/L	0.0014	0.0025	J	0.00096	0.005	03/31/2025
Sample ID: CCB2 Cyanide	mg/L	0.0013	0.0025	J	0.00096	0.005	03/31/2025
Sample ID: CCB3 Cyanide	mg/L	0.0015	0.0025	J	0.00096	0.005	03/31/2025

Initial and Continuing Calibration Blank Summary

Client: Weston Solutions, Inc.

SDG No.: Q1664

Project: RFP 905

RunNo.: LB135243

Analyte	Units	Result	Acceptance Limits	Conc Qual	MDL	RDL	Analysis Date
Sample ID: ICB1 Cyanide	mg/L	0.0012	0.0025	J	0.00096	0.005	03/31/2025
Sample ID: CCB1 Cyanide	mg/L	0.0015	0.0025	J	0.00096	0.005	03/31/2025
Sample ID: CCB2 Cyanide	mg/L	0.0018	0.0025	J	0.00096	0.005	03/31/2025
Sample ID: CCB3 Cyanide	mg/L	0.0015	0.0025	J	0.00096	0.005	03/31/2025