

Laboratory Control Sample Summary

Client: Weston Solutions, Inc.

SDG No.: Q1664

Project: RFP 905

Run No.: LB135233

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	PB167387BS							
Cyanide	mg/Kg	5	4.70		94	1	85-115	03/31/2025

Laboratory Control Sample Summary

Client: Weston Solutions, Inc.

SDG No.: Q1664

Project: RFP 905

Run No.: LB135243

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	PB167401BS							
Cyanide	mg/L	0.1	0.093		93	1	85-115	03/31/2025