


LB135232

7471B

INSTRUMENT ID: CV1

Linear



A= 0.0000e+000

B= 1.9269e-004

C= -5.5560e-002

Rho= 0.9998880

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	0.003	0.003	304	0.000	304					1
0.2	0.200	0.216	0.016	1407	0.0 %	1407					8
2.5	2.500	2.451	-0.049	13008	0.0 %	13008					-2
5.0	5.000	4.987	-0.013	26171	0.0 %	26171					0
7.5	7.500	7.607	0.107	39766	0.0 %	39766					1
10.0	10.000	9.936	-0.064	51855	0.0 %	51855					1

LB135232 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 31 Mar 2025 11:55:01

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	304	-	0.0000	7471B	PPB	31 Mar 2025 11:55:24	S	Std
0.2 - 1	50-2	1407	-	0.2000	7471B	PPB	31 Mar 2025 12:00:33	S	Std
2.5 - 1	52.5	13008	-	2.5000	7471B	PPB	31 Mar 2025 12:02:50	S	Std
5.0 - 1	55	26171	-	5.0000	7471B	PPB	31 Mar 2025 12:05:07	S	Std
7.5 - 1	57.5	39766	-	7.5000	7471B	PPB	31 Mar 2025 12:07:26	S	Std
10.0 - 1	510	51855	-	10.0000	7471B	PPB	31 Mar 2025 12:09:45	S	Std
ICV67 - 1	ICV67	22188	4.2198	-	7471B	PPB	31 Mar 2025 12:15:12	U	SMPL
ICB67 - 1	ICB67	516	0.0439	-	7471B	PPB	31 Mar 2025 12:17:28	U	SMPL
CCV103 - 1	CCV103	26049	4.9638	-	7471B	PPB	31 Mar 2025 12:19:46	U	SMPL
CCB103 - 1	CCB103	571	0.0545	-	7471B	PPB	31 Mar 2025 12:22:02	U	SMPL
CRA - 1	CRA	1363	0.2071	-	7471B	PPB	31 Mar 2025 12:24:19	U	SMPL
HighStd - 1	HighStd	51462	9.8606	-	7471B	PPB	31 Mar 2025 12:26:35	U	SMPL
ChkStd - 1	ChkStd	35151	6.7176	-	7471B	PPB	31 Mar 2025 12:28:50	U	SMPL
PB167386BL - 1	PBS	418	0.0250	-	7471B	PPB	31 Mar 2025 12:31:15	U	SMPL
PB167386BS - 1	LCSS	18739	3.5552	-	7471B	PPB	31 Mar 2025 12:35:32	U	SMPL
Q1168-03 - 1	MDL-SOIL-03-QT1-2025	565	0.0533	-	7471B	PPB	31 Mar 2025 12:37:47	U	SMPL
Q1664-01 - 1	P001-BBDGA-001-01	942	0.1260	-	7471B	PPB	31 Mar 2025 12:45:16	U	SMPL
Q1664-01DUP - 1	P001-BBDGA-001-01DUP	908	0.1194	-	7471B	PPB	31 Mar 2025 12:47:33	U	SMPL
Q1664-02 - 1	Q1664-01MS	23351	4.4439	-	7471B	PPB	31 Mar 2025 12:49:49	U	SMPL
Q1664-03 - 1	Q1664-01MSD	24526	4.6703	-	7471B	PPB	31 Mar 2025 12:52:05	U	SMPL
CCV104 - 1	CCV104	25298	4.8191	-	7471B	PPB	31 Mar 2025 12:54:22	U	SMPL
CCB104 - 1	CCB104	464	0.0338	-	7471B	PPB	31 Mar 2025 12:56:46	U	SMPL
Q1664-07 - 1	P001-BBDGA-001-02	2551	0.4360	-	7471B	PPB	31 Mar 2025 12:59:05	U	SMPL
Q1664-09 - 1	P001-BBDGA-002-01	574	0.0550	-	7471B	PPB	31 Mar 2025 13:01:20	U	SMPL
Q1664-11 - 1	P001-BBDGA-003-01	740	0.0870	-	7471B	PPB	31 Mar 2025 13:03:38	U	SMPL
Q1664-13 - 1	P001-BBDGA-004-01	768	0.0924	-	7471B	PPB	31 Mar 2025 13:05:54	U	SMPL
Q1664-15 - 1	P001-BBDGA-005-01	1004	0.1379	-	7471B	PPB	31 Mar 2025 13:08:10	U	SMPL
Q1664-17 - 1	P001-BBDGA-006-01	655	0.0707	-	7471B	PPB	31 Mar 2025 13:10:26	U	SMPL
Q1664-19 - 1	P001-BBDGA-007-01	847	0.1076	-	7471B	PPB	31 Mar 2025 13:12:43	U	SMPL
Q1664-21 - 1	P001-BBDGA-008-01	800	0.0986	-	7471B	PPB	31 Mar 2025 13:15:00	U	SMPL
Q1671-01 - 1	WC-1	13869	2.6168	-	7471B	PPB	31 Mar 2025 13:17:17	U	SMPL
Q1672-01 - 1	TP-8	1508	0.2350	-	7471B	PPB	31 Mar 2025 13:19:33	U	SMPL
CCV105 - 1	CCV105	26049	4.9638	-	7471B	PPB	31 Mar 2025 13:21:51	U	SMPL
CCB105 - 1	CCB105	545	0.0495	-	7471B	PPB	31 Mar 2025 13:24:06	U	SMPL
Q1674-01 - 1	RT5358	2922	0.5075	-	7471B	PPB	31 Mar 2025 13:26:24	U	SMPL
Q1674-03 - 1	72-11991	2549	0.4356	-	7471B	PPB	31 Mar 2025 13:28:40	U	SMPL
Q1678-01 - 1	MOO-25-0082	-44	-0.0640	-	7471B	PPB	31 Mar 2025 13:30:56	U	SMPL
Q1678-02 - 1	MOO-25-0086-87-88	1755	0.2826	-	7471B	PPB	31 Mar 2025 13:33:12	U	SMPL
Q1168-01 - 1		513	0.0433	-	7471B	PPB	31 Mar 2025 13:35:33	U	SMPL
Q1168-01RE - 1		1206	0.1768	-	7471B	PPB	31 Mar 2025 13:37:49	U	SMPL
Q1664-01LX5 - 1		152	-0.0263	-	7471B	PPB	31 Mar 2025 13:39:16	U	SMPL
Q1664-01A - 1		23960	4.5613	-	7471B	PPB	31 Mar 2025 13:41:33	U	SMPL
CCV106 - 1	CCV106	25869	4.9291	-	7471B	PPB	31 Mar 2025 13:43:49	U	SMPL
CCB106 - 1	CCB106	461	0.0333	-	7471B	PPB	31 Mar 2025 13:46:06	U	SMPL
							31 Mar 2025 14:00:41	U	SMPL