



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 2b -

CRDL STANDARD FOR AA & ICP

Client: Weston Solutions, Inc. **SDG No.:** Q1664
Contract: ROYF02 **Lab Code:** CHEM **Case No.:** Q1664 **SAS No.:** Q1664
Initial Calibration Source: _____
Continuing Calibration Source: _____

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CRA	Mercury	0.21	0.2	104	40 - 160	CV	03/31/2025	12:24	LB135232
CRI01	Aluminum	101	100	101	40 - 160	P	03/31/2025	13:24	LB135245
	Antimony	53.7	50.0	107	40 - 160	P	03/31/2025	13:24	LB135245
	Arsenic	21.6	20.0	108	40 - 160	P	03/31/2025	13:24	LB135245
	Barium	93.6	100	94	40 - 160	P	03/31/2025	13:24	LB135245
	Beryllium	5.94	6.0	99	40 - 160	P	03/31/2025	13:24	LB135245
	Cadmium	6.06	6.0	101	40 - 160	P	03/31/2025	13:24	LB135245
	Calcium	2010	2000	100	40 - 160	P	03/31/2025	13:24	LB135245
	Chromium	10.7	10.0	107	40 - 160	P	03/31/2025	13:24	LB135245
	Cobalt	29.9	30.0	100	40 - 160	P	03/31/2025	13:24	LB135245
	Copper	22.9	20.0	114	40 - 160	P	03/31/2025	13:24	LB135245
	Iron	87.0	100	87	40 - 160	P	03/31/2025	13:24	LB135245
	Lead	12.7	12.0	106	40 - 160	P	03/31/2025	13:24	LB135245
	Magnesium	2090	2000	104	40 - 160	P	03/31/2025	13:24	LB135245
	Manganese	19.5	20.0	98	40 - 160	P	03/31/2025	13:24	LB135245
	Nickel	40.5	40.0	101	40 - 160	P	03/31/2025	13:24	LB135245
	Potassium	2030	2000	101	40 - 160	P	03/31/2025	13:24	LB135245
	Selenium	20.3	20.0	101	40 - 160	P	03/31/2025	13:24	LB135245
	Silver	11.2	10.0	112	40 - 160	P	03/31/2025	13:24	LB135245
	Sodium	1840	2000	92	40 - 160	P	03/31/2025	13:24	LB135245
	Thallium	40.6	40.0	102	40 - 160	P	03/31/2025	13:24	LB135245
	Vanadium	39.9	40.0	100	40 - 160	P	03/31/2025	13:24	LB135245
	Zinc	44.7	40.0	112	40 - 160	P	03/31/2025	13:24	LB135245