



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client:	Weston Solutions, Inc.				SDG No.:	Q1664				
Contract:	ROYF02	Lab Code:	CHEM		Case No.:	Q1664	SAS No.:	Q1664		
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number	
ICB67	Mercury	0.20	+/-0.20	U	0.20	CV	03/31/2025	12:17	LB135232	

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Weston Solutions, Inc.

SDG No.: Q1664

Contract: ROYF02

Lab Code: CHEM

Case No.: Q1664

SAS No.: Q1664

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB103	Mercury	0.20	+/-0.20	U	0.20	CV	03/31/2025	12:22	LB135232
CCB104	Mercury	0.20	+/-0.20	U	0.20	CV	03/31/2025	12:56	LB135232
CCB105	Mercury	0.20	+/-0.20	U	0.20	CV	03/31/2025	13:24	LB135232
CCB106	Mercury	0.20	+/-0.20	U	0.20	CV	03/31/2025	14:00	LB135232

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Weston Solutions, Inc. **SDG No.:** Q1664
Contract: ROYF02 **Lab Code:** CHEM **Case No.:** Q1664 **SAS No.:** Q1664

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Aluminum	100	+/-100	U	100	P	03/31/2025	13:20	LB135245
	Antimony	50.0	+/-50.0	U	50.0	P	03/31/2025	13:20	LB135245
	Arsenic	20.0	+/-20.0	U	20.0	P	03/31/2025	13:20	LB135245
	Barium	100	+/-100	U	100	P	03/31/2025	13:20	LB135245
	Beryllium	6.00	+/-6.00	U	6.00	P	03/31/2025	13:20	LB135245
	Cadmium	6.00	+/-6.00	U	6.00	P	03/31/2025	13:20	LB135245
	Calcium	2000	+/-2000	U	2000	P	03/31/2025	13:20	LB135245
	Chromium	10.0	+/-10.0	U	10.0	P	03/31/2025	13:20	LB135245
	Cobalt	30.0	+/-30.0	U	30.0	P	03/31/2025	13:20	LB135245
	Copper	20.0	+/-20.0	U	20.0	P	03/31/2025	13:20	LB135245
	Iron	100	+/-100	U	100	P	03/31/2025	13:20	LB135245
	Lead	12.0	+/-12.0	U	12.0	P	03/31/2025	13:20	LB135245
	Magnesium	2000	+/-2000	U	2000	P	03/31/2025	13:20	LB135245
	Manganese	6.19	+/-20.0	J	20.0	P	03/31/2025	13:20	LB135245
	Nickel	40.0	+/-40.0	U	40.0	P	03/31/2025	13:20	LB135245
	Potassium	2000	+/-2000	U	2000	P	03/31/2025	13:20	LB135245
	Selenium	20.0	+/-20.0	U	20.0	P	03/31/2025	13:20	LB135245
	Silver	10.0	+/-10.0	U	10.0	P	03/31/2025	13:20	LB135245
	Sodium	2000	+/-2000	U	2000	P	03/31/2025	13:20	LB135245
	Thallium	40.0	+/-40.0	U	40.0	P	03/31/2025	13:20	LB135245
	Vanadium	40.0	+/-40.0	U	40.0	P	03/31/2025	13:20	LB135245
	Zinc	40.0	+/-40.0	U	40.0	P	03/31/2025	13:20	LB135245

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Weston Solutions, Inc.

SDG No.: Q1664

Contract: ROYF02

Lab Code: CHEM

Case No.: Q1664

SAS No.: Q1664

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Aluminum	100	+/-100	U	100	P	03/31/2025	13:41	LB135245
	Antimony	50.0	+/-50.0	U	50.0	P	03/31/2025	13:41	LB135245
	Arsenic	20.0	+/-20.0	U	20.0	P	03/31/2025	13:41	LB135245
	Barium	100	+/-100	U	100	P	03/31/2025	13:41	LB135245
	Beryllium	6.00	+/-6.00	U	6.00	P	03/31/2025	13:41	LB135245
	Cadmium	6.00	+/-6.00	U	6.00	P	03/31/2025	13:41	LB135245
	Calcium	2000	+/-2000	U	2000	P	03/31/2025	13:41	LB135245
	Chromium	10.0	+/-10.0	U	10.0	P	03/31/2025	13:41	LB135245
	Cobalt	30.0	+/-30.0	U	30.0	P	03/31/2025	13:41	LB135245
	Copper	20.0	+/-20.0	U	20.0	P	03/31/2025	13:41	LB135245
	Iron	100	+/-100	U	100	P	03/31/2025	13:41	LB135245
	Lead	12.0	+/-12.0	U	12.0	P	03/31/2025	13:41	LB135245
	Magnesium	2000	+/-2000	U	2000	P	03/31/2025	13:41	LB135245
	Manganese	6.15	+/-20.0	J	20.0	P	03/31/2025	13:41	LB135245
	Nickel	40.0	+/-40.0	U	40.0	P	03/31/2025	13:41	LB135245
	Potassium	2000	+/-2000	U	2000	P	03/31/2025	13:41	LB135245
	Selenium	20.0	+/-20.0	U	20.0	P	03/31/2025	13:41	LB135245
	Silver	10.0	+/-10.0	U	10.0	P	03/31/2025	13:41	LB135245
	Sodium	2000	+/-2000	U	2000	P	03/31/2025	13:41	LB135245
	Thallium	40.0	+/-40.0	U	40.0	P	03/31/2025	13:41	LB135245
	Vanadium	40.0	+/-40.0	U	40.0	P	03/31/2025	13:41	LB135245
	Zinc	40.0	+/-40.0	U	40.0	P	03/31/2025	13:41	LB135245
CCB02	Aluminum	100	+/-100	U	100	P	03/31/2025	14:58	LB135245
	Antimony	50.0	+/-50.0	U	50.0	P	03/31/2025	14:58	LB135245
	Arsenic	20.0	+/-20.0	U	20.0	P	03/31/2025	14:58	LB135245
	Barium	100	+/-100	U	100	P	03/31/2025	14:58	LB135245
	Beryllium	6.00	+/-6.00	U	6.00	P	03/31/2025	14:58	LB135245
	Cadmium	6.00	+/-6.00	U	6.00	P	03/31/2025	14:58	LB135245
	Calcium	2000	+/-2000	U	2000	P	03/31/2025	14:58	LB135245
	Chromium	10.0	+/-10.0	U	10.0	P	03/31/2025	14:58	LB135245
	Cobalt	30.0	+/-30.0	U	30.0	P	03/31/2025	14:58	LB135245
	Copper	20.0	+/-20.0	U	20.0	P	03/31/2025	14:58	LB135245
	Iron	100	+/-100	U	100	P	03/31/2025	14:58	LB135245
	Lead	12.0	+/-12.0	U	12.0	P	03/31/2025	14:58	LB135245
	Magnesium	2000	+/-2000	U	2000	P	03/31/2025	14:58	LB135245
	Manganese	5.91	+/-20.0	J	20.0	P	03/31/2025	14:58	LB135245
	Nickel	40.0	+/-40.0	U	40.0	P	03/31/2025	14:58	LB135245
	Potassium	2000	+/-2000	U	2000	P	03/31/2025	14:58	LB135245
	Selenium	20.0	+/-20.0	U	20.0	P	03/31/2025	14:58	LB135245

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Weston Solutions, Inc.

SDG No.: Q1664

Contract: ROYF02

Lab Code: CHEM

Case No.: Q1664

SAS No.: Q1664

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB02	Silver	1.18	+/-10.0	J	10.0	P	03/31/2025	14:58	LB135245
	Sodium	2000	+/-2000	U	2000	P	03/31/2025	14:58	LB135245
	Thallium	40.0	+/-40.0	U	40.0	P	03/31/2025	14:58	LB135245
	Vanadium	40.0	+/-40.0	U	40.0	P	03/31/2025	14:58	LB135245
	Zinc	40.0	+/-40.0	U	40.0	P	03/31/2025	14:58	LB135245
CCB03	Aluminum	100	+/-100	U	100	P	03/31/2025	15:44	LB135245
	Antimony	50.0	+/-50.0	U	50.0	P	03/31/2025	15:44	LB135245
	Arsenic	20.0	+/-20.0	U	20.0	P	03/31/2025	15:44	LB135245
	Barium	100	+/-100	U	100	P	03/31/2025	15:44	LB135245
	Beryllium	6.00	+/-6.00	U	6.00	P	03/31/2025	15:44	LB135245
	Cadmium	6.00	+/-6.00	U	6.00	P	03/31/2025	15:44	LB135245
	Calcium	2000	+/-2000	U	2000	P	03/31/2025	15:44	LB135245
	Chromium	10.0	+/-10.0	U	10.0	P	03/31/2025	15:44	LB135245
	Cobalt	30.0	+/-30.0	U	30.0	P	03/31/2025	15:44	LB135245
	Copper	20.0	+/-20.0	U	20.0	P	03/31/2025	15:44	LB135245
	Iron	100	+/-100	U	100	P	03/31/2025	15:44	LB135245
	Lead	12.0	+/-12.0	U	12.0	P	03/31/2025	15:44	LB135245
	Magnesium	2000	+/-2000	U	2000	P	03/31/2025	15:44	LB135245
	Manganese	5.72	+/-20.0	J	20.0	P	03/31/2025	15:44	LB135245
	Nickel	40.0	+/-40.0	U	40.0	P	03/31/2025	15:44	LB135245
	Potassium	2000	+/-2000	U	2000	P	03/31/2025	15:44	LB135245
	Selenium	20.0	+/-20.0	U	20.0	P	03/31/2025	15:44	LB135245
	Silver	1.43	+/-10.0	J	10.0	P	03/31/2025	15:44	LB135245
	Sodium	2000	+/-2000	U	2000	P	03/31/2025	15:44	LB135245
	Thallium	40.0	+/-40.0	U	40.0	P	03/31/2025	15:44	LB135245
	Vanadium	40.0	+/-40.0	U	40.0	P	03/31/2025	15:44	LB135245
	Zinc	40.0	+/-40.0	U	40.0	P	03/31/2025	15:44	LB135245
CCB04	Aluminum	100	+/-100	U	100	P	03/31/2025	16:02	LB135245
	Antimony	50.0	+/-50.0	U	50.0	P	03/31/2025	16:02	LB135245
	Arsenic	20.0	+/-20.0	U	20.0	P	03/31/2025	16:02	LB135245
	Barium	100	+/-100	U	100	P	03/31/2025	16:02	LB135245
	Beryllium	6.00	+/-6.00	U	6.00	P	03/31/2025	16:02	LB135245
	Cadmium	6.00	+/-6.00	U	6.00	P	03/31/2025	16:02	LB135245
	Calcium	2000	+/-2000	U	2000	P	03/31/2025	16:02	LB135245
	Chromium	10.0	+/-10.0	U	10.0	P	03/31/2025	16:02	LB135245
	Cobalt	30.0	+/-30.0	U	30.0	P	03/31/2025	16:02	LB135245
	Copper	20.0	+/-20.0	U	20.0	P	03/31/2025	16:02	LB135245
	Iron	100	+/-100	U	100	P	03/31/2025	16:02	LB135245
	Lead	12.0	+/-12.0	U	12.0	P	03/31/2025	16:02	LB135245

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Weston Solutions, Inc. **SDG No.:** Q1664
Contract: ROYF02 **Lab Code:** CHEM **Case No.:** Q1664 **SAS No.:** Q1664

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB04	Magnesium	2000	+/-2000	U	2000	P	03/31/2025	16:02	LB135245
	Manganese	5.86	+/-20.0	J	20.0	P	03/31/2025	16:02	LB135245
	Nickel	40.0	+/-40.0	U	40.0	P	03/31/2025	16:02	LB135245
	Potassium	2000	+/-2000	U	2000	P	03/31/2025	16:02	LB135245
	Selenium	20.0	+/-20.0	U	20.0	P	03/31/2025	16:02	LB135245
	Silver	1.26	+/-10.0	J	10.0	P	03/31/2025	16:02	LB135245
	Sodium	2000	+/-2000	U	2000	P	03/31/2025	16:02	LB135245
	Thallium	40.0	+/-40.0	U	40.0	P	03/31/2025	16:02	LB135245
	Vanadium	40.0	+/-40.0	U	40.0	P	03/31/2025	16:02	LB135245
	Zinc	40.0	+/-40.0	U	40.0	P	03/31/2025	16:02	LB135245