

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Weston Solutions, Inc. **SDG No.:** Q1664
Contract: ROYF02 **Lab Code:** CHEM **Case No.:** Q1664 **SAS No.:** Q1664
Initial Calibration Source: EPA
Continuing Calibration Source: PLASMA-PURE

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV68	Mercury	3.92	4.0	98	90 - 110	CV	04/01/2025	10:51	LB135246

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Weston Solutions, Inc. **SDG No.:** Q1664
Contract: ROYF02 **Lab Code:** CHEM **Case No.:** Q1664 **SAS No.:** Q1664
Initial Calibration Source: EPA
Continuing Calibration Source: PLASMA-PURE

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV07	Mercury	4.99	5.0	100	90 - 110	CV	04/01/2025	10:59	LB135246
CCV08	Mercury	5.07	5.0	101	90 - 110	CV	04/01/2025	11:29	LB135246
CCV09	Mercury	4.81	5.0	96	90 - 110	CV	04/01/2025	12:01	LB135246
CCV10	Mercury	4.64	5.0	93	90 - 110	CV	04/01/2025	12:24	LB135246
CCV11	Mercury	4.87	5.0	97	90 - 110	CV	04/01/2025	13:02	LB135246

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Weston Solutions, Inc. **SDG No.:** Q1664

Contract: ROYF02 **Lab Code:** CHEM **Case No.:** Q1664 **SAS No.:** Q1664

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV01	Aluminum	2480	2500	99	90 - 110	P	04/01/2025	10:57	LB135255
	Antimony	1010	1000	101	90 - 110	P	04/01/2025	10:57	LB135255
	Arsenic	996	1000	100	90 - 110	P	04/01/2025	10:57	LB135255
	Barium	484	520	93	90 - 110	P	04/01/2025	10:57	LB135255
	Beryllium	491	510	96	90 - 110	P	04/01/2025	10:57	LB135255
	Cadmium	494	510	97	90 - 110	P	04/01/2025	10:57	LB135255
	Calcium	9700	10000	97	90 - 110	P	04/01/2025	10:57	LB135255
	Chromium	519	520	100	90 - 110	P	04/01/2025	10:57	LB135255
	Cobalt	502	520	96	90 - 110	P	04/01/2025	10:57	LB135255
	Copper	521	510	102	90 - 110	P	04/01/2025	10:57	LB135255
	Iron	9910	10000	99	90 - 110	P	04/01/2025	10:57	LB135255
	Lead	979	1000	98	90 - 110	P	04/01/2025	10:57	LB135255
	Magnesium	5690	6000	95	90 - 110	P	04/01/2025	10:57	LB135255
	Manganese	490	520	94	90 - 110	P	04/01/2025	10:57	LB135255
	Nickel	507	530	96	90 - 110	P	04/01/2025	10:57	LB135255
	Potassium	9620	9900	97	90 - 110	P	04/01/2025	10:57	LB135255
	Selenium	1030	1000	103	90 - 110	P	04/01/2025	10:57	LB135255
	Silver	252	250	101	90 - 110	P	04/01/2025	10:57	LB135255
	Sodium	9360	10000	94	90 - 110	P	04/01/2025	10:57	LB135255
	Thallium	1040	1000	104	90 - 110	P	04/01/2025	10:57	LB135255
	Vanadium	481	500	96	90 - 110	P	04/01/2025	10:57	LB135255
	Zinc	1010	1000	101	90 - 110	P	04/01/2025	10:57	LB135255

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Weston Solutions, Inc. **SDG No.:** Q1664
Contract: ROYF02 **Lab Code:** CHEM **Case No.:** Q1664 **SAS No.:** Q1664
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
LLICV01	Aluminum	97.1	100	97	80 - 120	P	04/01/2025	11:15	LB135255
	Antimony	53.4	50.0	107	80 - 120	P	04/01/2025	11:15	LB135255
	Arsenic	21.1	20.0	106	80 - 120	P	04/01/2025	11:15	LB135255
	Barium	85.7	100	86	80 - 120	P	04/01/2025	11:15	LB135255
	Beryllium	5.74	6.0	96	80 - 120	P	04/01/2025	11:15	LB135255
	Cadmium	5.75	6.0	96	80 - 120	P	04/01/2025	11:15	LB135255
	Calcium	1920	2000	96	80 - 120	P	04/01/2025	11:15	LB135255
	Chromium	9.96	10.0	100	80 - 120	P	04/01/2025	11:15	LB135255
	Cobalt	28.7	30.0	96	80 - 120	P	04/01/2025	11:15	LB135255
	Copper	22.6	20.0	113	80 - 120	P	04/01/2025	11:15	LB135255
	Iron	97.2	100	97	80 - 120	P	04/01/2025	11:15	LB135255
	Lead	11.7	12.0	98	80 - 120	P	04/01/2025	11:15	LB135255
	Magnesium	2000	2000	100	80 - 120	P	04/01/2025	11:15	LB135255
	Manganese	20.3	20.0	101	80 - 120	P	04/01/2025	11:15	LB135255
	Nickel	38.8	40.0	97	80 - 120	P	04/01/2025	11:15	LB135255
	Potassium	1830	2000	91	80 - 120	P	04/01/2025	11:15	LB135255
	Selenium	21.8	20.0	109	80 - 120	P	04/01/2025	11:15	LB135255
	Silver	10.6	10.0	106	80 - 120	P	04/01/2025	11:15	LB135255
	Sodium	1770	2000	89	80 - 120	P	04/01/2025	11:15	LB135255
	Thallium	40.6	40.0	102	80 - 120	P	04/01/2025	11:15	LB135255
	Vanadium	38.5	40.0	96	80 - 120	P	04/01/2025	11:15	LB135255
	Zinc	44.4	40.0	111	80 - 120	P	04/01/2025	11:15	LB135255

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Weston Solutions, Inc. **SDG No.:** Q1664

Contract: ROYF02 **Lab Code:** CHEM **Case No.:** Q1664 **SAS No.:** Q1664

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV01	Aluminum	9840	10000	98	90 - 110	P	04/01/2025	11:50	LB135255
	Antimony	5100	5000	102	90 - 110	P	04/01/2025	11:50	LB135255
	Arsenic	5030	5000	101	90 - 110	P	04/01/2025	11:50	LB135255
	Barium	9720	10000	97	90 - 110	P	04/01/2025	11:50	LB135255
	Beryllium	243	250	97	90 - 110	P	04/01/2025	11:50	LB135255
	Cadmium	2440	2500	98	90 - 110	P	04/01/2025	11:50	LB135255
	Calcium	23900	25000	96	90 - 110	P	04/01/2025	11:50	LB135255
	Chromium	1010	1000	101	90 - 110	P	04/01/2025	11:50	LB135255
	Cobalt	2440	2500	98	90 - 110	P	04/01/2025	11:50	LB135255
	Copper	1260	1250	101	90 - 110	P	04/01/2025	11:50	LB135255
	Iron	4890	5000	98	90 - 110	P	04/01/2025	11:50	LB135255
	Lead	4950	5000	99	90 - 110	P	04/01/2025	11:50	LB135255
	Magnesium	23600	25000	94	90 - 110	P	04/01/2025	11:50	LB135255
	Manganese	2390	2500	95	90 - 110	P	04/01/2025	11:50	LB135255
	Nickel	2460	2500	98	90 - 110	P	04/01/2025	11:50	LB135255
	Potassium	24900	25000	100	90 - 110	P	04/01/2025	11:50	LB135255
	Selenium	5090	5000	102	90 - 110	P	04/01/2025	11:50	LB135255
	Silver	1260	1250	100	90 - 110	P	04/01/2025	11:50	LB135255
	Sodium	24500	25000	98	90 - 110	P	04/01/2025	11:50	LB135255
	Thallium	5110	5000	102	90 - 110	P	04/01/2025	11:50	LB135255
	Vanadium	2440	2500	98	90 - 110	P	04/01/2025	11:50	LB135255
	Zinc	2540	2500	102	90 - 110	P	04/01/2025	11:50	LB135255
CCV02	Aluminum	9790	10000	98	90 - 110	P	04/01/2025	12:43	LB135255
	Antimony	5200	5000	104	90 - 110	P	04/01/2025	12:43	LB135255
	Arsenic	5090	5000	102	90 - 110	P	04/01/2025	12:43	LB135255
	Barium	9360	10000	94	90 - 110	P	04/01/2025	12:43	LB135255
	Beryllium	238	250	95	90 - 110	P	04/01/2025	12:43	LB135255
	Cadmium	2420	2500	97	90 - 110	P	04/01/2025	12:43	LB135255
	Calcium	23600	25000	94	90 - 110	P	04/01/2025	12:43	LB135255
	Chromium	1010	1000	101	90 - 110	P	04/01/2025	12:43	LB135255
	Cobalt	2410	2500	96	90 - 110	P	04/01/2025	12:43	LB135255
	Copper	1250	1250	100	90 - 110	P	04/01/2025	12:43	LB135255
	Iron	4850	5000	97	90 - 110	P	04/01/2025	12:43	LB135255
	Lead	4950	5000	99	90 - 110	P	04/01/2025	12:43	LB135255

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Weston Solutions, Inc. **SDG No.:** Q1664

Contract: ROYF02 **Lab Code:** CHEM **Case No.:** Q1664 **SAS No.:** Q1664

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV02	Magnesium	23100	25000	92	90 - 110	P	04/01/2025	12:43	LB135255
	Manganese	2340	2500	94	90 - 110	P	04/01/2025	12:43	LB135255
	Nickel	2460	2500	98	90 - 110	P	04/01/2025	12:43	LB135255
	Potassium	25100	25000	100	90 - 110	P	04/01/2025	12:43	LB135255
	Selenium	5180	5000	104	90 - 110	P	04/01/2025	12:43	LB135255
	Silver	1240	1250	99	90 - 110	P	04/01/2025	12:43	LB135255
	Sodium	24100	25000	96	90 - 110	P	04/01/2025	12:43	LB135255
	Thallium	5110	5000	102	90 - 110	P	04/01/2025	12:43	LB135255
	Vanadium	2410	2500	96	90 - 110	P	04/01/2025	12:43	LB135255
	Zinc	2520	2500	101	90 - 110	P	04/01/2025	12:43	LB135255
CCV03	Aluminum	10000	10000	100	90 - 110	P	04/01/2025	13:35	LB135255
	Antimony	5390	5000	108	90 - 110	P	04/01/2025	13:35	LB135255
	Arsenic	5260	5000	105	90 - 110	P	04/01/2025	13:35	LB135255
	Barium	9580	10000	96	90 - 110	P	04/01/2025	13:35	LB135255
	Beryllium	242	250	97	90 - 110	P	04/01/2025	13:35	LB135255
	Cadmium	2460	2500	98	90 - 110	P	04/01/2025	13:35	LB135255
	Calcium	23800	25000	95	90 - 110	P	04/01/2025	13:35	LB135255
	Chromium	1020	1000	102	90 - 110	P	04/01/2025	13:35	LB135255
	Cobalt	2450	2500	98	90 - 110	P	04/01/2025	13:35	LB135255
	Copper	1290	1250	103	90 - 110	P	04/01/2025	13:35	LB135255
	Iron	4960	5000	99	90 - 110	P	04/01/2025	13:35	LB135255
	Lead	5050	5000	101	90 - 110	P	04/01/2025	13:35	LB135255
	Magnesium	23100	25000	92	90 - 110	P	04/01/2025	13:35	LB135255
	Manganese	2570	2500	103	90 - 110	P	04/01/2025	13:35	LB135255
	Nickel	2510	2500	100	90 - 110	P	04/01/2025	13:35	LB135255
	Potassium	25300	25000	101	90 - 110	P	04/01/2025	13:35	LB135255
	Selenium	5390	5000	108	90 - 110	P	04/01/2025	13:35	LB135255
	Silver	1270	1250	102	90 - 110	P	04/01/2025	13:35	LB135255
	Sodium	24600	25000	98	90 - 110	P	04/01/2025	13:35	LB135255
	Thallium	5300	5000	106	90 - 110	P	04/01/2025	13:35	LB135255
	Vanadium	2460	2500	98	90 - 110	P	04/01/2025	13:35	LB135255
	Zinc	2570	2500	103	90 - 110	P	04/01/2025	13:35	LB135255
CCV04	Aluminum	9860	10000	99	90 - 110	P	04/01/2025	14:31	LB135255
	Antimony	5310	5000	106	90 - 110	P	04/01/2025	14:31	LB135255

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Weston Solutions, Inc. SDG No.: Q1664
Contract: ROYF02 Lab Code: CHEM Case No.: Q1664 SAS No.: Q1664
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV04	Arsenic	5180	5000	104	90 - 110	P	04/01/2025	14:31	LB135255
	Barium	9490	10000	95	90 - 110	P	04/01/2025	14:31	LB135255
	Beryllium	242	250	97	90 - 110	P	04/01/2025	14:31	LB135255
	Cadmium	2440	2500	98	90 - 110	P	04/01/2025	14:31	LB135255
	Calcium	23700	25000	95	90 - 110	P	04/01/2025	14:31	LB135255
	Chromium	1020	1000	102	90 - 110	P	04/01/2025	14:31	LB135255
	Cobalt	2430	2500	97	90 - 110	P	04/01/2025	14:31	LB135255
	Copper	1270	1250	102	90 - 110	P	04/01/2025	14:31	LB135255
	Iron	4860	5000	97	90 - 110	P	04/01/2025	14:31	LB135255
	Lead	5020	5000	100	90 - 110	P	04/01/2025	14:31	LB135255
	Magnesium	22900	25000	92	90 - 110	P	04/01/2025	14:31	LB135255
	Manganese	2560	2500	102	90 - 110	P	04/01/2025	14:31	LB135255
	Nickel	2490	2500	100	90 - 110	P	04/01/2025	14:31	LB135255
	Potassium	24900	25000	100	90 - 110	P	04/01/2025	14:31	LB135255
	Selenium	5300	5000	106	90 - 110	P	04/01/2025	14:31	LB135255
	Silver	1260	1250	100	90 - 110	P	04/01/2025	14:31	LB135255
	Sodium	24500	25000	98	90 - 110	P	04/01/2025	14:31	LB135255
	Thallium	5210	5000	104	90 - 110	P	04/01/2025	14:31	LB135255
	Vanadium	2440	2500	98	90 - 110	P	04/01/2025	14:31	LB135255
	Zinc	2540	2500	102	90 - 110	P	04/01/2025	14:31	LB135255
CCV05	Aluminum	9760	10000	98	90 - 110	P	04/01/2025	15:23	LB135255
	Antimony	5200	5000	104	90 - 110	P	04/01/2025	15:23	LB135255
	Arsenic	5090	5000	102	90 - 110	P	04/01/2025	15:23	LB135255
	Barium	9500	10000	95	90 - 110	P	04/01/2025	15:23	LB135255
	Beryllium	241	250	96	90 - 110	P	04/01/2025	15:23	LB135255
	Cadmium	2420	2500	97	90 - 110	P	04/01/2025	15:23	LB135255
	Calcium	23600	25000	94	90 - 110	P	04/01/2025	15:23	LB135255
	Chromium	995	1000	100	90 - 110	P	04/01/2025	15:23	LB135255
	Cobalt	2410	2500	96	90 - 110	P	04/01/2025	15:23	LB135255
	Copper	1250	1250	100	90 - 110	P	04/01/2025	15:23	LB135255
	Iron	4710	5000	94	90 - 110	P	04/01/2025	15:23	LB135255
	Lead	4970	5000	99	90 - 110	P	04/01/2025	15:23	LB135255
	Magnesium	22800	25000	91	90 - 110	P	04/01/2025	15:23	LB135255
	Manganese	2540	2500	102	90 - 110	P	04/01/2025	15:23	LB135255

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Weston Solutions, Inc. **SDG No.:** Q1664

Contract: ROYF02 **Lab Code:** CHEM **Case No.:** Q1664 **SAS No.:** Q1664

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV05	Nickel	2460	2500	98	90 - 110	P	04/01/2025	15:23	LB135255
	Potassium	24000	25000	96	90 - 110	P	04/01/2025	15:23	LB135255
	Selenium	5210	5000	104	90 - 110	P	04/01/2025	15:23	LB135255
	Silver	1220	1250	98	90 - 110	P	04/01/2025	15:23	LB135255
	Sodium	23300	25000	93	90 - 110	P	04/01/2025	15:23	LB135255
	Thallium	5130	5000	103	90 - 110	P	04/01/2025	15:23	LB135255
	Vanadium	2420	2500	97	90 - 110	P	04/01/2025	15:23	LB135255
	Zinc	2470	2500	99	90 - 110	P	04/01/2025	15:23	LB135255
CCV06	Aluminum	9810	10000	98	90 - 110	P	04/01/2025	15:52	LB135255
	Antimony	5170	5000	104	90 - 110	P	04/01/2025	15:52	LB135255
	Arsenic	5070	5000	101	90 - 110	P	04/01/2025	15:52	LB135255
	Barium	9450	10000	94	90 - 110	P	04/01/2025	15:52	LB135255
	Beryllium	242	250	97	90 - 110	P	04/01/2025	15:52	LB135255
	Cadmium	2420	2500	97	90 - 110	P	04/01/2025	15:52	LB135255
	Calcium	23700	25000	95	90 - 110	P	04/01/2025	15:52	LB135255
	Chromium	1000	1000	100	90 - 110	P	04/01/2025	15:52	LB135255
	Cobalt	2400	2500	96	90 - 110	P	04/01/2025	15:52	LB135255
	Copper	1250	1250	100	90 - 110	P	04/01/2025	15:52	LB135255
	Iron	4740	5000	95	90 - 110	P	04/01/2025	15:52	LB135255
	Lead	4960	5000	99	90 - 110	P	04/01/2025	15:52	LB135255
	Magnesium	22700	25000	91	90 - 110	P	04/01/2025	15:52	LB135255
	Manganese	2540	2500	101	90 - 110	P	04/01/2025	15:52	LB135255
	Nickel	2460	2500	98	90 - 110	P	04/01/2025	15:52	LB135255
	Potassium	24000	25000	96	90 - 110	P	04/01/2025	15:52	LB135255
	Selenium	5180	5000	104	90 - 110	P	04/01/2025	15:52	LB135255
	Silver	1230	1250	98	90 - 110	P	04/01/2025	15:52	LB135255
	Sodium	23500	25000	94	90 - 110	P	04/01/2025	15:52	LB135255
	Thallium	5120	5000	102	90 - 110	P	04/01/2025	15:52	LB135255
	Vanadium	2410	2500	96	90 - 110	P	04/01/2025	15:52	LB135255
	Zinc	2480	2500	99	90 - 110	P	04/01/2025	15:52	LB135255
CCV07	Aluminum	9820	10000	98	90 - 110	P	04/01/2025	16:56	LB135255
	Antimony	5170	5000	103	90 - 110	P	04/01/2025	16:56	LB135255
	Arsenic	5060	5000	101	90 - 110	P	04/01/2025	16:56	LB135255
	Barium	9420	10000	94	90 - 110	P	04/01/2025	16:56	LB135255

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Weston Solutions, Inc. **SDG No.:** Q1664

Contract: ROYF02 **Lab Code:** CHEM **Case No.:** Q1664 **SAS No.:** Q1664

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV07	Beryllium	248	250	99	90 - 110	P	04/01/2025	16:56	LB135255
	Cadmium	2420	2500	97	90 - 110	P	04/01/2025	16:56	LB135255
	Calcium	23600	25000	95	90 - 110	P	04/01/2025	16:56	LB135255
	Chromium	1010	1000	101	90 - 110	P	04/01/2025	16:56	LB135255
	Cobalt	2400	2500	96	90 - 110	P	04/01/2025	16:56	LB135255
	Copper	1250	1250	100	90 - 110	P	04/01/2025	16:56	LB135255
	Iron	4800	5000	96	90 - 110	P	04/01/2025	16:56	LB135255
	Lead	4970	5000	99	90 - 110	P	04/01/2025	16:56	LB135255
	Magnesium	22900	25000	92	90 - 110	P	04/01/2025	16:56	LB135255
	Manganese	2570	2500	103	90 - 110	P	04/01/2025	16:56	LB135255
	Nickel	2470	2500	99	90 - 110	P	04/01/2025	16:56	LB135255
	Potassium	24100	25000	96	90 - 110	P	04/01/2025	16:56	LB135255
	Selenium	5160	5000	103	90 - 110	P	04/01/2025	16:56	LB135255
	Silver	1250	1250	100	90 - 110	P	04/01/2025	16:56	LB135255
	Sodium	23700	25000	95	90 - 110	P	04/01/2025	16:56	LB135255
	Thallium	5010	5000	100	90 - 110	P	04/01/2025	16:56	LB135255
	Vanadium	2410	2500	96	90 - 110	P	04/01/2025	16:56	LB135255
	Zinc	2520	2500	101	90 - 110	P	04/01/2025	16:56	LB135255
CCV08	Aluminum	9900	10000	99	90 - 110	P	04/01/2025	17:41	LB135255
	Antimony	5280	5000	106	90 - 110	P	04/01/2025	17:41	LB135255
	Arsenic	5170	5000	103	90 - 110	P	04/01/2025	17:41	LB135255
	Barium	9630	10000	96	90 - 110	P	04/01/2025	17:41	LB135255
	Beryllium	241	250	96	90 - 110	P	04/01/2025	17:41	LB135255
	Cadmium	2460	2500	98	90 - 110	P	04/01/2025	17:41	LB135255
	Calcium	23800	25000	95	90 - 110	P	04/01/2025	17:41	LB135255
	Chromium	1020	1000	102	90 - 110	P	04/01/2025	17:41	LB135255
	Cobalt	2440	2500	98	90 - 110	P	04/01/2025	17:41	LB135255
	Copper	1290	1250	103	90 - 110	P	04/01/2025	17:41	LB135255
	Iron	4940	5000	99	90 - 110	P	04/01/2025	17:41	LB135255
	Lead	5040	5000	101	90 - 110	P	04/01/2025	17:41	LB135255
	Magnesium	22800	25000	91	90 - 110	P	04/01/2025	17:41	LB135255
	Manganese	2260	2500	90	90 - 110	P	04/01/2025	17:41	LB135255
	Nickel	2500	2500	100	90 - 110	P	04/01/2025	17:41	LB135255
	Potassium	25400	25000	102	90 - 110	P	04/01/2025	17:41	LB135255

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Weston Solutions, Inc. **SDG No.:** Q1664
Contract: ROYF02 **Lab Code:** CHEM **Case No.:** Q1664 **SAS No.:** Q1664
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV08	Selenium	5280	5000	106	90 - 110	P	04/01/2025	17:41	LB135255
	Silver	1260	1250	101	90 - 110	P	04/01/2025	17:41	LB135255
	Sodium	25000	25000	100	90 - 110	P	04/01/2025	17:41	LB135255
	Thallium	5110	5000	102	90 - 110	P	04/01/2025	17:41	LB135255
	Vanadium	2440	2500	98	90 - 110	P	04/01/2025	17:41	LB135255
	Zinc	2530	2500	101	90 - 110	P	04/01/2025	17:41	LB135255