



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 2b -

CRDL STANDARD FOR AA & ICP

Client: Weston Solutions, Inc. **SDG No.:** Q1664
Contract: ROYF02 **Lab Code:** CHEM **Case No.:** Q1664 **SAS No.:** Q1664
Initial Calibration Source: _____
Continuing Calibration Source: _____

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CRA	Mercury	0.24	0.2	118	40 - 160	CV	04/01/2025	11:03	LB135246
CRI01	Aluminum	94.5	100	94	40 - 160	P	04/01/2025	11:28	LB135255
	Antimony	52.0	50.0	104	40 - 160	P	04/01/2025	11:28	LB135255
	Arsenic	20.3	20.0	101	40 - 160	P	04/01/2025	11:28	LB135255
	Barium	87.8	100	88	40 - 160	P	04/01/2025	11:28	LB135255
	Beryllium	5.78	6.0	96	40 - 160	P	04/01/2025	11:28	LB135255
	Cadmium	5.60	6.0	93	40 - 160	P	04/01/2025	11:28	LB135255
	Calcium	1950	2000	98	40 - 160	P	04/01/2025	11:28	LB135255
	Chromium	9.89	10.0	99	40 - 160	P	04/01/2025	11:28	LB135255
	Cobalt	28.5	30.0	95	40 - 160	P	04/01/2025	11:28	LB135255
	Copper	22.6	20.0	113	40 - 160	P	04/01/2025	11:28	LB135255
	Iron	100	100	100	40 - 160	P	04/01/2025	11:28	LB135255
	Lead	11.7	12.0	98	40 - 160	P	04/01/2025	11:28	LB135255
	Magnesium	2030	2000	102	40 - 160	P	04/01/2025	11:28	LB135255
	Manganese	20.8	20.0	104	40 - 160	P	04/01/2025	11:28	LB135255
	Nickel	38.5	40.0	96	40 - 160	P	04/01/2025	11:28	LB135255
	Potassium	1780	2000	89	40 - 160	P	04/01/2025	11:28	LB135255
	Selenium	19.7	20.0	98	40 - 160	P	04/01/2025	11:28	LB135255
	Silver	10.3	10.0	103	40 - 160	P	04/01/2025	11:28	LB135255
	Sodium	1760	2000	88	40 - 160	P	04/01/2025	11:28	LB135255
	Thallium	41.2	40.0	103	40 - 160	P	04/01/2025	11:28	LB135255
	Vanadium	37.8	40.0	94	40 - 160	P	04/01/2025	11:28	LB135255
	Zinc	43.7	40.0	109	40 - 160	P	04/01/2025	11:28	LB135255