

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: Weston Solutions, Inc. **SDG No.:** Q1664
Contract: ROYF02 **Lab Code:** CHEM **Case No.:** Q1664 **SAS No.:** Q1664

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB167392BS							
Aluminum	ug/L	1000	948		95	80 - 120	P
Antimony	ug/L	400	411		103	80 - 120	P
Arsenic	ug/L	400	388		97	80 - 120	P
Barium	ug/L	100	85.6		86	80 - 120	P
Beryllium	ug/L	100	96.8		97	80 - 120	P
Cadmium	ug/L	100	93.4		93	80 - 120	P
Calcium	ug/L	500	487	J	97	80 - 120	P
Chromium	ug/L	200	203		102	80 - 120	P
Cobalt	ug/L	100	94.0		94	80 - 120	P
Copper	ug/L	150	153		102	80 - 120	P
Iron	ug/L	1500	1390		93	80 - 120	P
Lead	ug/L	500	475		95	80 - 120	P
Magnesium	ug/L	1000	883	J	88	80 - 120	P
Manganese	ug/L	100	102		102	80 - 120	P
Nickel	ug/L	250	242		97	80 - 120	P
Potassium	ug/L	5000	4520		90	80 - 120	P
Selenium	ug/L	1000	999		100	80 - 120	P
Silver	ug/L	37.5	37.3		100	80 - 120	P
Sodium	ug/L	1500	1300		87	80 - 120	P
Thallium	ug/L	1000	987		99	80 - 120	P
Vanadium	ug/L	150	140		93	80 - 120	P
Zinc	ug/L	100	100		100	80 - 120	P

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: Weston Solutions, Inc.

SDG No.: Q1664

Contract: ROYF02

Lab Code: CHEM

Case No.: Q1664

SAS No.: Q1664

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB167415BS Mercury	ug/L	4.0	3.75		94	80 - 120	CV