

Continuing Calibration Report for SequenceID : FC032825AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FC068432.D

Aliphatic C9-C12	9984223.000	60.000	3.100	6.360	166403.717	162319.116	-2.516
Aliphatic C12-C16	6826421.000	40.000	6.361	9.742	170660.525	168194.168	-1.466
Aliphatic C16-C21	10083173.000	60.000	9.743	13.094	168052.883	171510.591	2.016
Aliphatic C21-C28	12341117.000	80.000	13.095	16.746	154263.963	168869.942	8.649
Aliphatic C28-C40	14271477.000	120.000	16.747	21.538	118928.975	136171.748	12.663
Aliphatic EPH	53506411.000	360.000	3.100	21.538	148628.919	157243.762	5.479