

Continuing Calibration Report for SequenceID : FF032825AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FF015742.D

Aliphatic C9-C12	7590297.000	60.000	3.211	6.833	126504.950	128465.799	1.526
Aliphatic C12-C16	5306204.000	40.000	6.834	10.268	132655.100	131614.301	-0.791
Aliphatic C16-C21	8259169.000	60.000	10.269	13.634	137652.817	135374.394	-1.683
Aliphatic C21-C28	11178024.000	80.000	13.635	17.200	139725.300	132278.233	-5.630
Aliphatic C28-C40	13103539.000	120.000	17.201	22.202	109196.158	108147.044	-0.970
Aliphatic EPH	45437233.000	360.000	3.211	22.202	126214.536	124041.354	-1.752