

Continuing Calibration Report for SequenceID : FF032825AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FF015752.D

Aliphatic C9-C12	7088622.000	60.000	3.211	6.833	118143.700	128465.799	8.035
Aliphatic C12-C16	4911388.000	40.000	6.834	10.268	122784.700	131614.301	6.709
Aliphatic C16-C21	7679869.000	60.000	10.269	13.634	127997.817	135374.394	5.449
Aliphatic C21-C28	10007110.000	80.000	13.635	17.200	125088.875	132278.233	5.435
Aliphatic C28-C40	11793922.000	120.000	17.201	22.202	98282.683	108147.044	9.121
Aliphatic EPH	41480911.000	360.000	3.211	22.202	115224.753	124041.354	7.108