

Continuing Calibration Report for SequenceID : FG032825AR

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FG015572.D

Aromatic C10-C12	5212196.000	40.000	4.449	6.329	130304.900	133996.317	2.755
Aromatic C12-C16	8020452.000	60.000	6.330	9.039	133674.200	137151.728	2.536
Aromatic C16-C21	10337835.000	80.000	9.040	13.371	129222.938	139410.441	7.308
Aromatic C21-C36	21669615.000	180.000	13.372	18.837	120386.750	130193.774	7.533
Aromatic EPH	45240098.000	360.000	4.449	18.837	125666.939	133824.086	6.095