

Continuing Calibration Report for SequenceID : FG032825AR

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FG015585.D

Aromatic C10-C12	5174568.000	40.000	4.449	6.329	129364.200	133996.317	3.457
Aromatic C12-C16	8034793.000	60.000	6.330	9.039	133913.217	137151.728	2.361
Aromatic C16-C21	10315765.000	80.000	9.040	13.371	128947.063	139410.441	7.505
Aromatic C21-C36	21978345.000	180.000	13.372	18.837	122101.917	130193.774	6.215
Aromatic EPH	45503471.000	360.000	4.449	18.837	126398.531	133824.086	5.549