

Continuing Calibration Report for SequenceID : FG032825AR

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FG015591.D

Aromatic C10-C12	5178861.000	40.000	4.449	6.329	129471.525	133996.317	3.377
Aromatic C12-C16	8038714.000	60.000	6.330	9.039	133978.567	137151.728	2.314
Aromatic C16-C21	10318195.000	80.000	9.040	13.371	128977.438	139410.441	7.484
Aromatic C21-C36	21839978.000	180.000	13.372	18.837	121333.211	130193.774	6.806
Aromatic EPH	45375748.000	360.000	4.449	18.837	126043.744	133824.086	5.814