

Continuing Calibration Report for SequenceID : FG040125AR

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FG015594.D

Aromatic C10-C12	5474581.000	40.000	4.450	6.329	136864.525	133996.317	-2.141
Aromatic C12-C16	8515969.000	60.000	6.330	9.035	141932.817	137151.728	-3.486
Aromatic C16-C21	10771263.000	80.000	9.036	13.365	134640.788	139410.441	3.421
Aromatic C21-C36	22584191.000	180.000	13.366	18.829	125467.728	130193.774	3.630
Aromatic EPH	47346004.000	360.000	4.450	18.829	131516.678	133824.086	1.724