

WATER EPH MATRIX SPIKE/MATRIX SPIKE DUPLICATE RECOVERY

Lab Name: Chemtech **Client:** Weston Solutions, Inc.
Lab Code: CHEM **Cas No:** Q1664 **SAS No :** Q1664 **SDG No:** Q1664
Sample No : Q1664-05 **Datafile:** FC068473.D
Client ID : P001-BBDGA-001-01MS

COMPOUND	SPIKE ADDED ug/l	SAMPLE CONCENTRATION ug/l	MS/MSD CONCENTRATION ug/l	% REC	Qual	QC LIMITS
Aliphatic C9-C12	300	10.1	296	95		(40-140)
Aliphatic C12-C16	200	32.9	300	133		(40-140)
Aliphatic C16-C21	300	13.6	338	108		(40-140)
Aliphatic C21-C28	400	0	375	94		(40-140)
Aliphatic C28-C40	600	39.1	566	88		(40-140)