

WATER EPH MATRIX SPIKE/MATRIX SPIKE DUPLICATE RECOVERY

Lab Name: Chemtech **Client:** Weston Solutions, Inc.
Lab Code: CHEM **Cas No:** Q1664 **SAS No :** Q1664 **SDG No:** Q1664
Sample No : Q1664-06 **Datafile:** FC068474.D
Client ID : P001-BBDGA-001-01MSD

COMPOUND	SPIKE ADDED ug/l	SAMPLE CONCENTRATION ug/l	MS/MSD CONCENTRATION ug/l	% REC	Qual	RPD QC LIMITS QC Limit Of RPD
Aliphatic C9-C12	300	10.1	302	97		2.08 (40-140) 25
Aliphatic C12-C16	200	32.9	304	135		1.49 (40-140) 25
Aliphatic C16-C21	300	13.6	342	109		0.92 (40-140) 25
Aliphatic C21-C28	400	0	379	95		1.06 (40-140) 25
Aliphatic C28-C40	600	39.1	572	89		1.13 (40-140) 25