

**SOLID EPH LABORATORY CONTROL SPIKE/LABORATORY CONTROL SPIKE DUPLICATE RECOVERY**

**Lab Name:** Chemtech **Client:** Weston Solutions, Inc.  
**Lab Code:** CHEM **Cas No:** Q1664 **SAS No :** Q1664 **SDG No:** Q1664  
**Sample No :** PB167362BSD **Datafile:** FG015575.D  
**Client ID :** PB167362BSD

| COMPOUND         | SPIKE<br>ADDED<br>mg/kg | LCS/LCSD<br>CONCENTRATION<br>mg/kg | % REC | Qual | RPD   QC LIMITS | QC Limit<br>Of RPD |
|------------------|-------------------------|------------------------------------|-------|------|-----------------|--------------------|
| Aromatic C10-C12 | 6.7                     | 4.49                               | 67    |      | 6.4  (40-140)   | 50                 |
| Aromatic C12-C16 | 10                      | 7.88                               | 79    |      | 6.5  (40-140)   | 50                 |
| Aromatic C16-C21 | 16.7                    | 15.0                               | 90    |      | 6.8  (40-140)   | 50                 |
| Aromatic C21-C36 | 26.6                    | 26.1                               | 98    |      | 6.7  (40-140)   | 50                 |