

WATER EPH LABORATORY CONTROL SPIKE/LABORATORY CONTROL SPIKE DUPLICATE RECOVERY

Lab Name: Chemtech **Client:** Weston Solutions, Inc.
Lab Code: CHEM **Cas No:** Q1664 **SAS No :** Q1664 **SDG No:** Q1664
Sample No : PB167402BS **Datafile:** FG015596.D
Client ID : PB167402BS

COMPOUND	SPIKE ADDED ug/l	LCS/LCSD CONCENTRATION ug/l	% REC	Qual	QC LIMITS
Aromatic C10-C12	200	161	80		(40-140)
Aromatic C12-C16	300	319	106		(40-140)
Aromatic C16-C21	500	582	116		(40-140)
Aromatic C21-C36	800	935	116		(40-140)