

CALIBRATION VERIFICATION SUMMARY

Contract: ROYF02

Lab Code: CHEM Case No.: Q1664 SAS No.: Q1664 SDG NO.: Q1664

Continuing Calib Date: 03/28/2025 Initial Calibration Date(s): 03/18/2025 03/18/2025

Continuing Calib Time: 16:13 Initial Calibration Time(s): 14:03 22:15

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1242-1	(1)	4.78	4.79	4.69	4.89	0.01
Aroclor-1242-2	(2)	4.80	4.81	4.71	4.91	0.01
Aroclor-1242-3	(3)	4.86	4.86	4.76	4.96	0.00
Aroclor-1242-4	(4)	4.98	4.98	4.88	5.08	0.00
Aroclor-1242-5	(5)	5.63	5.64	5.54	5.74	0.01
Tetrachloro-m-xylene		3.69	3.69	3.59	3.79	0.00
Decachlorobiphenyl		8.74	8.74	8.64	8.84	0.00

CALIBRATION VERIFICATION SUMMARY

Contract: ROYF02

Lab Code: CHEM Case No.: Q1664 SAS No.: Q1664 SDG NO.: Q1664

Continuing Calib Date: 03/28/2025 Initial Calibration Date(s): 03/18/2025 03/18/2025

Continuing Calib Time: 16:13 Initial Calibration Time(s): 14:03 22:15

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1242-1	(1)	4.77	4.77	4.67	4.87	0.00
Aroclor-1242-2	(2)	4.79	4.79	4.69	4.89	0.00
Aroclor-1242-3	(3)	4.96	4.97	4.87	5.07	0.01
Aroclor-1242-4	(4)	5.05	5.05	4.95	5.15	0.00
Aroclor-1242-5	(5)	5.57	5.57	5.47	5.67	0.00
Tetrachloro-m-xylene		3.69	3.69	3.59	3.79	0.00
Decachlorobiphenyl		8.69	8.70	8.60	8.80	0.01