

### CALIBRATION VERIFICATION SUMMARY

Contract: ROYF02

Lab Code: CHEM Case No.: Q1664 SAS No.: Q1664 SDG NO.: Q1664

Continuing Calib Date: 03/28/2025 Initial Calibration Date(s): 03/27/2025 03/27/2025

Continuing Calib Time: 08:29 Initial Calibration Time(s): 10:30 17:53

GC Column: ZB-MR1 ID: 0.32 (mm)

| COMPOUND             |     | CCAL<br>RT | AVG<br>RT | RT WINDOW |       | DIFF<br>RT |
|----------------------|-----|------------|-----------|-----------|-------|------------|
|                      |     |            |           | FROM      | TO    |            |
| Aroclor-1016-1       | (1) | 5.67       | 5.66      | 5.56      | 5.76  | -0.01      |
| Aroclor-1016-2       | (2) | 5.69       | 5.69      | 5.59      | 5.79  | 0.00       |
| Aroclor-1016-3       | (3) | 5.76       | 5.75      | 5.65      | 5.85  | -0.01      |
| Aroclor-1016-4       | (4) | 5.85       | 5.85      | 5.75      | 5.95  | 0.00       |
| Aroclor-1016-5       | (5) | 6.15       | 6.14      | 6.04      | 6.24  | 0.00       |
| Aroclor-1260-1       | (1) | 7.26       | 7.26      | 7.16      | 7.36  | 0.00       |
| Aroclor-1260-2       | (2) | 7.52       | 7.51      | 7.41      | 7.61  | -0.01      |
| Aroclor-1260-3       | (3) | 7.88       | 7.87      | 7.77      | 7.97  | -0.01      |
| Aroclor-1260-4       | (4) | 8.10       | 8.09      | 7.99      | 8.19  | -0.01      |
| Aroclor-1260-5       | (5) | 8.42       | 8.41      | 8.31      | 8.51  | -0.01      |
| Tetrachloro-m-xylene |     | 4.52       | 4.51      | 4.41      | 4.61  | -0.01      |
| Decachlorobiphenyl   |     | 10.24      | 10.23     | 10.13     | 10.33 | -0.01      |

### CALIBRATION VERIFICATION SUMMARY

Contract: ROYF02

Lab Code: CHEM Case No.: Q1664 SAS No.: Q1664 SDG NO.: Q1664

Continuing Calib Date: 03/28/2025 Initial Calibration Date(s): 03/27/2025 03/27/2025

Continuing Calib Time: 08:29 Initial Calibration Time(s): 10:30 17:53

GC Column: ZB-MR2 ID: 0.32 (mm)

| COMPOUND             |     | CCAL<br>RT | AVG<br>RT | RT WINDOW<br>FROM   TO |      | DIFF<br>RT |
|----------------------|-----|------------|-----------|------------------------|------|------------|
| Aroclor-1016-1       | (1) | 4.90       | 4.90      | 4.80                   | 5.00 | 0.00       |
| Aroclor-1016-2       | (2) | 4.92       | 4.92      | 4.82                   | 5.02 | 0.00       |
| Aroclor-1016-3       | (3) | 5.10       | 5.10      | 5.00                   | 5.20 | 0.00       |
| Aroclor-1016-4       | (4) | 5.14       | 5.14      | 5.04                   | 5.24 | 0.00       |
| Aroclor-1016-5       | (5) | 5.36       | 5.36      | 5.26                   | 5.46 | 0.00       |
| Aroclor-1260-1       | (1) | 6.39       | 6.39      | 6.29                   | 6.49 | 0.00       |
| Aroclor-1260-2       | (2) | 6.58       | 6.58      | 6.48                   | 6.68 | 0.00       |
| Aroclor-1260-3       | (3) | 6.73       | 6.73      | 6.63                   | 6.83 | 0.00       |
| Aroclor-1260-4       | (4) | 7.21       | 7.21      | 7.11                   | 7.31 | 0.00       |
| Aroclor-1260-5       | (5) | 7.45       | 7.45      | 7.35                   | 7.55 | 0.00       |
| Tetrachloro-m-xylene |     | 3.82       | 3.82      | 3.72                   | 3.92 | 0.00       |
| Decachlorobiphenyl   |     | 8.86       | 8.86      | 8.76                   | 8.96 | 0.00       |