

CALIBRATION VERIFICATION SUMMARY

Contract: ROYF02

Lab Code: CHEM Case No.: Q1664 SAS No.: Q1664 SDG NO.: Q1664

Continuing Calib Date: 03/28/2025 Initial Calibration Date(s): 03/27/2025 03/27/2025

Continuing Calib Time: 19:09 Initial Calibration Time(s): 10:30 17:53

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1242-1	(1)	5.68	5.66	5.56	5.76	-0.02
Aroclor-1242-2	(2)	5.70	5.69	5.59	5.79	-0.01
Aroclor-1242-3	(3)	5.76	5.75	5.65	5.85	-0.01
Aroclor-1242-4	(4)	5.86	5.85	5.75	5.95	-0.01
Aroclor-1242-5	(5)	6.59	6.58	6.48	6.68	-0.01
Tetrachloro-m-xylene		4.52	4.51	4.41	4.61	-0.01
Decachlorobiphenyl		10.25	10.23	10.13	10.33	-0.02

CALIBRATION VERIFICATION SUMMARY

Contract: ROYF02

Lab Code: CHEM Case No.: Q1664 SAS No.: Q1664 SDG NO.: Q1664

Continuing Calib Date: 03/28/2025 Initial Calibration Date(s): 03/27/2025 03/27/2025

Continuing Calib Time: 19:09 Initial Calibration Time(s): 10:30 17:53

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1242-1	(1)	4.90	4.90	4.80	5.00	0.00
Aroclor-1242-2	(2)	4.92	4.92	4.82	5.02	0.00
Aroclor-1242-3	(3)	5.10	5.10	5.00	5.20	0.00
Aroclor-1242-4	(4)	5.19	5.18	5.08	5.28	0.00
Aroclor-1242-5	(5)	5.71	5.71	5.61	5.81	0.00
Tetrachloro-m-xylene		3.82	3.82	3.72	3.92	0.00
Decachlorobiphenyl		8.86	8.86	8.76	8.96	0.00