

CALIBRATION VERIFICATION SUMMARY

Contract: ROYF02

Lab Code: CHEM Case No.: Q1664 SAS No.: Q1664 SDG NO.: Q1664

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 03/18/2025 03/18/2025

Client Sample No.: CCAL02 Date Analyzed: 03/28/2025

Lab Sample No.: AR1660CCC500 Data File : PO110098.D Time Analyzed: 15:35

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.783	4.685	4.885	485.000	500.000	-3.0
Aroclor-1016-2	4.802	4.704	4.904	484.280	500.000	-3.1
Aroclor-1016-3	4.858	4.761	4.961	484.340	500.000	-3.1
Aroclor-1016-4	4.979	4.881	5.081	481.750	500.000	-3.7
Aroclor-1016-5	5.236	5.139	5.339	531.300	500.000	6.3
Aroclor-1260-1	6.276	6.180	6.380	480.190	500.000	-4.0
Aroclor-1260-2	6.465	6.369	6.569	467.130	500.000	-6.6
Aroclor-1260-3	6.833	6.737	6.937	463.480	500.000	-7.3
Aroclor-1260-4	7.093	6.998	7.198	443.830	500.000	-11.2
Aroclor-1260-5	7.335	7.239	7.439	430.320	500.000	-13.9
Decachlorobiphenyl	8.739	8.645	8.845	60.610	50.000	21.2
Tetrachloro-m-xylene	3.691	3.592	3.792	48.110	50.000	-3.8

CALIBRATION VERIFICATION SUMMARY

Contract: ROYF02

Lab Code: CHEM Case No.: Q1664 SAS No.: Q1664 SDG NO.: Q1664

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 03/18/2025 03/18/2025

Client Sample No.: CCAL02 Date Analyzed: 03/28/2025

Lab Sample No.: AR1660CCC500 Data File : PO110098.D Time Analyzed: 15:35

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.770	4.672	4.872	484.520	500.000	-3.1
Aroclor-1016-2	4.789	4.691	4.891	479.650	500.000	-4.1
Aroclor-1016-3	4.964	4.867	5.067	482.010	500.000	-3.6
Aroclor-1016-4	5.007	4.909	5.109	429.650	500.000	-14.1
Aroclor-1016-5	5.219	5.122	5.322	527.830	500.000	5.6
Aroclor-1260-1	6.251	6.155	6.355	462.490	500.000	-7.5
Aroclor-1260-2	6.438	6.342	6.542	457.840	500.000	-8.4
Aroclor-1260-3	6.591	6.495	6.695	436.250	500.000	-12.8
Aroclor-1260-4	7.063	6.967	7.167	429.410	500.000	-14.1
Aroclor-1260-5	7.303	7.207	7.407	423.970	500.000	-15.2
Decachlorobiphenyl	8.691	8.596	8.796	55.050	50.000	10.1
Tetrachloro-m-xylene	3.689	3.590	3.790	49.940	50.000	-0.1