

CALIBRATION VERIFICATION SUMMARY

Contract: ROYF02

Lab Code: CHEM Case No.: Q1664 SAS No.: Q1664 SDG NO.: Q1664

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 03/27/2025 03/27/2025

Client Sample No.: CCAL03 Date Analyzed: 03/28/2025

Lab Sample No.: AR1660CCC500 Data File : PP070966.D Time Analyzed: 13:22

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.676	5.563	5.763	498.650	500.000	-0.3
Aroclor-1016-2	5.696	5.585	5.785	450.210	500.000	-10.0
Aroclor-1016-3	5.760	5.648	5.848	477.180	500.000	-4.6
Aroclor-1016-4	5.858	5.745	5.945	510.330	500.000	2.1
Aroclor-1016-5	6.151	6.038	6.238	520.720	500.000	4.1
Aroclor-1260-1	7.269	7.156	7.356	518.440	500.000	3.7
Aroclor-1260-2	7.524	7.410	7.610	530.170	500.000	6.0
Aroclor-1260-3	7.882	7.768	7.968	507.260	500.000	1.5
Aroclor-1260-4	8.107	7.992	8.192	491.880	500.000	-1.6
Aroclor-1260-5	8.427	8.311	8.511	519.460	500.000	3.9
Decachlorobiphenyl	10.247	10.125	10.325	53.330	50.000	6.7
Tetrachloro-m-xylene	4.522	4.412	4.612	50.810	50.000	1.6

CALIBRATION VERIFICATION SUMMARY

Contract: ROYF02

Lab Code: CHEM Case No.: Q1664 SAS No.: Q1664 SDG NO.: Q1664

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 03/27/2025 03/27/2025

Client Sample No.: CCAL03 Date Analyzed: 03/28/2025

Lab Sample No.: AR1660CCC500 Data File : PP070966.D Time Analyzed: 13:22

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.905	4.803	5.003	467.000	500.000	-6.6
Aroclor-1016-2	4.924	4.822	5.022	451.750	500.000	-9.7
Aroclor-1016-3	5.101	4.999	5.199	474.970	500.000	-5.0
Aroclor-1016-4	5.142	5.041	5.241	491.520	500.000	-1.7
Aroclor-1016-5	5.358	5.255	5.455	487.690	500.000	-2.5
Aroclor-1260-1	6.393	6.292	6.492	496.620	500.000	-0.7
Aroclor-1260-2	6.581	6.480	6.680	506.510	500.000	1.3
Aroclor-1260-3	6.735	6.634	6.834	512.990	500.000	2.6
Aroclor-1260-4	7.206	7.105	7.305	494.610	500.000	-1.1
Aroclor-1260-5	7.447	7.346	7.546	467.970	500.000	-6.4
Decachlorobiphenyl	8.860	8.759	8.959	52.120	50.000	4.2
Tetrachloro-m-xylene	3.818	3.716	3.916	50.330	50.000	0.7