

CALIBRATION VERIFICATION SUMMARY

Contract: ROYF02

Lab Code: CHEM Case No.: Q1664 SAS No.: Q1664 SDG NO.: Q1664

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 03/27/2025 03/27/2025

Client Sample No.: CCAL04 Date Analyzed: 03/28/2025

Lab Sample No.: AR1660CCC500 Data File : PP070981.D Time Analyzed: 18:20

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.674	5.563	5.763	513.150	500.000	2.6
Aroclor-1016-2	5.696	5.585	5.785	499.340	500.000	-0.1
Aroclor-1016-3	5.758	5.648	5.848	495.660	500.000	-0.9
Aroclor-1016-4	5.855	5.745	5.945	533.360	500.000	6.7
Aroclor-1016-5	6.148	6.038	6.238	538.240	500.000	7.6
Aroclor-1260-1	7.268	7.156	7.356	546.070	500.000	9.2
Aroclor-1260-2	7.522	7.410	7.610	550.390	500.000	10.1
Aroclor-1260-3	7.881	7.768	7.968	538.090	500.000	7.6
Aroclor-1260-4	8.105	7.992	8.192	516.920	500.000	3.4
Aroclor-1260-5	8.426	8.311	8.511	540.310	500.000	8.1
Decachlorobiphenyl	10.246	10.125	10.325	54.710	50.000	9.4
Tetrachloro-m-xylene	4.519	4.412	4.612	52.100	50.000	4.2

CALIBRATION VERIFICATION SUMMARY

Contract: ROYF02

Lab Code: CHEM Case No.: Q1664 SAS No.: Q1664 SDG NO.: Q1664

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 03/27/2025 03/27/2025

Client Sample No.: CCAL04 Date Analyzed: 03/28/2025

Lab Sample No.: AR1660CCC500 Data File : PP070981.D Time Analyzed: 18:20

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.904	4.803	5.003	505.820	500.000	1.2
Aroclor-1016-2	4.923	4.822	5.022	488.840	500.000	-2.2
Aroclor-1016-3	5.100	4.999	5.199	523.890	500.000	4.8
Aroclor-1016-4	5.142	5.041	5.241	535.980	500.000	7.2
Aroclor-1016-5	5.357	5.255	5.455	515.150	500.000	3.0
Aroclor-1260-1	6.393	6.292	6.492	531.560	500.000	6.3
Aroclor-1260-2	6.581	6.480	6.680	534.260	500.000	6.9
Aroclor-1260-3	6.734	6.634	6.834	546.610	500.000	9.3
Aroclor-1260-4	7.206	7.105	7.305	523.050	500.000	4.6
Aroclor-1260-5	7.447	7.346	7.546	482.900	500.000	-3.4
Decachlorobiphenyl	8.861	8.759	8.959	50.220	50.000	0.4
Tetrachloro-m-xylene	3.817	3.716	3.916	50.950	50.000	1.9