

NEW JERSEY LAB ID#:20012 : NEW YORK LAB ID#: 11376

GC ANALYSIS CONFORMANCE/NON-CONFORMANCE SUMMARY

CHEMTECH PROJECT NUMBER: PL031125

	NA	NO	YES
1. Chromatograms Labeled/Compounds Identified	✓		
2. PEM Meet Criteria.	✓		
3. Calibration - Initial Calibration performed within 30 days before sample analysis and continuing calibration performed within 24 hours of sample analysis, 12 HOURS IF 8000 SERIES METHOD.	✓		
3(a).Initial calibration Meet Criteria If not met, list those compounds and their recoveries which fall outside the acceptable range. COMMENTS:	✓		
3(b).Continuous calibration(CCC) Meet Criteria If not met, list those compounds and their recoveries which fall outside the acceptable range. COMMENTS:	✓		
4. Blank Contamination - If yes, list compounds and concentrations in each blank: COMMENTS:	✓		
5. Surrogate Recoveries Meet Criteria If not met, list those compounds and their recoveries which fall outside the acceptable ranges. COMMENTS: Q1502-11 -- TCMX high in 2nd column Q1502-09 -- TCMX high in 1st column			✓

6(a).Matrix Spike/Matrix Spike Duplicate Recoveries Meet Criteria

✓

If not met, list those compounds and their recoveries which fall outside the acceptable

COMMENTS:

6(b).Blank Spike/Blank Spike Duplicate Recoveries Meet Criteria

✓

If not met, list those compounds and their recoveries which fall outside the acceptable

COMMENTS:

PB167077BS -- Comp#2 recovery fail

7. Retention Time Shift Meet Criteria (if applicable)

✓

COMMENTS:

8. Extraction Holding Time Met

✓

If not met, list number of days exceeded for each sample:

COMMENTS:

9. Analysis Holding Time Met

✓

If not met, list those compounds and their recoveries which fall outside the acceptable range.

COMMENTS:

ADDITIONAL COMMENTS:

Dhvanil

QA REVIEW

03/12/2025

Date