

**CALIBRATION VERIFICATION SUMMARY**

**Contract:** ROYF02

**Lab Code:** CHEM      **Case No.:** Q1664      **SAS No.:** Q1664      **SDG NO.:** Q1664

**Continuing Calib Date:** 04/01/2025      **Initial Calibration Date(s):** 03/18/2025      03/18/2025

**Continuing Calib Time:** 09:17      **Initial Calibration Time(s):** 14:03      22:15

**GC Column:** ZB-MR1      **ID:** 0.32      (mm)

| COMPOUND             | CCAL<br>RT | AVG<br>RT | RT WINDOW |      | DIFF<br>RT |
|----------------------|------------|-----------|-----------|------|------------|
|                      |            |           | FROM      | TO   |            |
| Tetrachloro-m-xylene | 3.69       | 3.69      | 3.59      | 3.79 | 0.00       |
| Decachlorobiphenyl   | 8.74       | 8.75      | 8.65      | 8.85 | 0.01       |

**CALIBRATION VERIFICATION SUMMARY**

**Contract:** ROYF02

**Lab Code:** CHEM      **Case No.:** Q1664      **SAS No.:** Q1664      **SDG NO.:** Q1664

**Continuing Calib Date:** 04/01/2025      **Initial Calibration Date(s):** 03/18/2025      03/18/2025

**Continuing Calib Time:** 09:17      **Initial Calibration Time(s):** 14:03      22:15

**GC Column:** ZB-MR2      **ID:** 0.32 (mm)

| COMPOUND             | CCAL<br>RT | AVG<br>RT | RT WINDOW |      | DIFF<br>RT |
|----------------------|------------|-----------|-----------|------|------------|
|                      |            |           | FROM      | TO   |            |
| Tetrachloro-m-xylene | 3.69       | 3.69      | 3.59      | 3.79 | 0.00       |
| Decachlorobiphenyl   | 8.69       | 8.70      | 8.60      | 8.80 | 0.01       |