

CALIBRATION VERIFICATION SUMMARY

Contract: ROYF02

Lab Code: CHEM Case No.: Q1664 SAS No.: Q1664 SDG NO.: Q1664

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 03/18/2025 03/18/2025

Client Sample No.: CCAL01 Date Analyzed: 03/31/2025

Lab Sample No.: AR1660CCC500 Data File : PO110118.D Time Analyzed: 16:29

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.785	4.685	4.885	538.480	500.000	7.7
Aroclor-1016-2	4.804	4.704	4.904	548.250	500.000	9.7
Aroclor-1016-3	4.860	4.761	4.961	546.100	500.000	9.2
Aroclor-1016-4	4.981	4.881	5.081	546.370	500.000	9.3
Aroclor-1016-5	5.239	5.139	5.339	543.190	500.000	8.6
Aroclor-1260-1	6.279	6.180	6.380	544.600	500.000	8.9
Aroclor-1260-2	6.468	6.369	6.569	524.610	500.000	4.9
Aroclor-1260-3	6.836	6.737	6.937	528.470	500.000	5.7
Aroclor-1260-4	7.097	6.998	7.198	517.950	500.000	3.6
Aroclor-1260-5	7.338	7.239	7.439	515.420	500.000	3.1
Decachlorobiphenyl	8.744	8.645	8.845	53.850	50.000	7.7
Tetrachloro-m-xylene	3.692	3.592	3.792	53.800	50.000	7.6

CALIBRATION VERIFICATION SUMMARY

Contract: ROYF02

Lab Code: CHEM Case No.: Q1664 SAS No.: Q1664 SDG NO.: Q1664

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 03/18/2025 03/18/2025

Client Sample No.: CCAL01 Date Analyzed: 03/31/2025

Lab Sample No.: AR1660CCC500 Data File : PO110118.D Time Analyzed: 16:29

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.771	4.672	4.872	534.260	500.000	6.9
Aroclor-1016-2	4.790	4.691	4.891	530.240	500.000	6.0
Aroclor-1016-3	4.966	4.867	5.067	535.940	500.000	7.2
Aroclor-1016-4	5.008	4.909	5.109	529.170	500.000	5.8
Aroclor-1016-5	5.221	5.122	5.322	536.180	500.000	7.2
Aroclor-1260-1	6.253	6.155	6.355	512.810	500.000	2.6
Aroclor-1260-2	6.441	6.342	6.542	504.090	500.000	0.8
Aroclor-1260-3	6.594	6.495	6.695	490.210	500.000	-2.0
Aroclor-1260-4	7.066	6.967	7.167	485.320	500.000	-2.9
Aroclor-1260-5	7.306	7.207	7.407	481.080	500.000	-3.8
Decachlorobiphenyl	8.695	8.596	8.796	48.510	50.000	-3.0
Tetrachloro-m-xylene	3.689	3.590	3.790	54.630	50.000	9.3