

CALIBRATION VERIFICATION SUMMARY

Contract: ROYF02

Lab Code: CHEM Case No.: Q1664 SAS No.: Q1664 SDG NO.: Q1664

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 03/18/2025 03/18/2025

Client Sample No.: CCAL02 Date Analyzed: 03/31/2025

Lab Sample No.: AR1660CCC500 Data File : PO110133.D Time Analyzed: 21:45

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.784	4.685	4.885	538.380	500.000	7.7
Aroclor-1016-2	4.803	4.704	4.904	547.160	500.000	9.4
Aroclor-1016-3	4.859	4.761	4.961	547.030	500.000	9.4
Aroclor-1016-4	4.980	4.881	5.081	545.170	500.000	9.0
Aroclor-1016-5	5.238	5.139	5.339	557.730	500.000	11.5
Aroclor-1260-1	6.278	6.180	6.380	550.050	500.000	10.0
Aroclor-1260-2	6.468	6.369	6.569	531.560	500.000	6.3
Aroclor-1260-3	6.835	6.737	6.937	541.170	500.000	8.2
Aroclor-1260-4	7.096	6.998	7.198	523.210	500.000	4.6
Aroclor-1260-5	7.338	7.239	7.439	528.670	500.000	5.7
Decachlorobiphenyl	8.743	8.645	8.845	58.730	50.000	17.5
Tetrachloro-m-xylene	3.691	3.592	3.792	53.340	50.000	6.7

CALIBRATION VERIFICATION SUMMARY

Contract: ROYF02

Lab Code: CHEM Case No.: Q1664 SAS No.: Q1664 SDG NO.: Q1664

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 03/18/2025 03/18/2025

Client Sample No.: CCAL02 Date Analyzed: 03/31/2025

Lab Sample No.: AR1660CCC500 Data File : PO110133.D Time Analyzed: 21:45

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	4.770	4.672	4.872	534.000	500.000	6.8
Aroclor-1016-2	4.790	4.691	4.891	533.650	500.000	6.7
Aroclor-1016-3	4.965	4.867	5.067	534.730	500.000	6.9
Aroclor-1016-4	5.007	4.909	5.109	517.030	500.000	3.4
Aroclor-1016-5	5.220	5.122	5.322	552.200	500.000	10.4
Aroclor-1260-1	6.252	6.155	6.355	517.650	500.000	3.5
Aroclor-1260-2	6.440	6.342	6.542	513.240	500.000	2.6
Aroclor-1260-3	6.593	6.495	6.695	498.640	500.000	-0.3
Aroclor-1260-4	7.065	6.967	7.167	496.040	500.000	-0.8
Aroclor-1260-5	7.306	7.207	7.407	496.630	500.000	-0.7
Decachlorobiphenyl	8.695	8.596	8.796	51.890	50.000	3.8
Tetrachloro-m-xylene	3.689	3.590	3.790	54.440	50.000	8.9