

### CALIBRATION VERIFICATION SUMMARY

Contract: ROYF02

Lab Code: CHEM Case No.: Q1664 SAS No.: Q1664 SDG NO.: Q1664

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 03/18/2025 03/18/2025

Client Sample No.: CCAL03 Date Analyzed: 04/01/2025

Lab Sample No.: AR1660CCC500 Data File : PO110139.D Time Analyzed: 08:58

| COMPOUND             | RT    | RT WINDOW |       | CALC<br>AMOUNT(ng) | NOM<br>AMOUNT(ng) | %D   |
|----------------------|-------|-----------|-------|--------------------|-------------------|------|
|                      |       | FROM      | TO    |                    |                   |      |
| Aroclor-1016-1       | 4.786 | 4.685     | 4.885 | 477.190            | 500.000           | -4.6 |
| Aroclor-1016-2       | 4.805 | 4.704     | 4.904 | 477.640            | 500.000           | -4.5 |
| Aroclor-1016-3       | 4.861 | 4.761     | 4.961 | 484.730            | 500.000           | -3.1 |
| Aroclor-1016-4       | 4.982 | 4.881     | 5.081 | 482.680            | 500.000           | -3.5 |
| Aroclor-1016-5       | 5.240 | 5.139     | 5.339 | 533.930            | 500.000           | 6.8  |
| Aroclor-1260-1       | 6.280 | 6.180     | 6.380 | 493.670            | 500.000           | -1.3 |
| Aroclor-1260-2       | 6.469 | 6.369     | 6.569 | 467.720            | 500.000           | -6.5 |
| Aroclor-1260-3       | 6.837 | 6.737     | 6.937 | 473.610            | 500.000           | -5.3 |
| Aroclor-1260-4       | 7.097 | 6.998     | 7.198 | 466.460            | 500.000           | -6.7 |
| Aroclor-1260-5       | 7.340 | 7.239     | 7.439 | 466.860            | 500.000           | -6.6 |
| Decachlorobiphenyl   | 8.744 | 8.645     | 8.845 | 53.010             | 50.000            | 6.0  |
| Tetrachloro-m-xylene | 3.693 | 3.592     | 3.792 | 47.200             | 50.000            | -5.6 |

### CALIBRATION VERIFICATION SUMMARY

Contract: ROYF02

Lab Code: CHEM Case No.: Q1664 SAS No.: Q1664 SDG NO.: Q1664

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 03/18/2025 03/18/2025

Client Sample No.: CCAL03 Date Analyzed: 04/01/2025

Lab Sample No.: AR1660CCC500 Data File : PO110139.D Time Analyzed: 08:58

| COMPOUND             | RT    | RT WINDOW |       | CALC<br>AMOUNT(ng) | NOM<br>AMOUNT(ng) | %D    |
|----------------------|-------|-----------|-------|--------------------|-------------------|-------|
|                      |       | FROM      | TO    |                    |                   |       |
| Aroclor-1016-1       | 4.769 | 4.672     | 4.872 | 454.360            | 500.000           | -9.1  |
| Aroclor-1016-2       | 4.788 | 4.691     | 4.891 | 451.710            | 500.000           | -9.7  |
| Aroclor-1016-3       | 4.964 | 4.867     | 5.067 | 454.170            | 500.000           | -9.2  |
| Aroclor-1016-4       | 5.006 | 4.909     | 5.109 | 400.000            | 500.000           | -20.0 |
| Aroclor-1016-5       | 5.219 | 5.122     | 5.322 | 499.100            | 500.000           | -0.2  |
| Aroclor-1260-1       | 6.252 | 6.155     | 6.355 | 446.280            | 500.000           | -10.7 |
| Aroclor-1260-2       | 6.439 | 6.342     | 6.542 | 442.830            | 500.000           | -11.4 |
| Aroclor-1260-3       | 6.592 | 6.495     | 6.695 | 425.630            | 500.000           | -14.9 |
| Aroclor-1260-4       | 7.063 | 6.967     | 7.167 | 428.480            | 500.000           | -14.3 |
| Aroclor-1260-5       | 7.304 | 7.207     | 7.407 | 425.940            | 500.000           | -14.8 |
| Decachlorobiphenyl   | 8.693 | 8.596     | 8.796 | 45.510             | 50.000            | -9.0  |
| Tetrachloro-m-xylene | 3.687 | 3.590     | 3.790 | 45.990             | 50.000            | -8.0  |