

CALIBRATION VERIFICATION SUMMARY

Contract: ROYF02

Lab Code: CHEM Case No.: Q1664 SAS No.: Q1664 SDG NO.: Q1664

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 03/18/2025 03/18/2025

Client Sample No.: CCAL04 Date Analyzed: 04/01/2025

Lab Sample No.: AR1660CCC500 Data File : PO110147.D Time Analyzed: 13:47

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.783	4.685	4.885	473.110	500.000	-5.4
Aroclor-1016-2	4.802	4.704	4.904	484.240	500.000	-3.2
Aroclor-1016-3	4.859	4.761	4.961	485.860	500.000	-2.8
Aroclor-1016-4	4.979	4.881	5.081	484.320	500.000	-3.1
Aroclor-1016-5	5.237	5.139	5.339	539.090	500.000	7.8
Aroclor-1260-1	6.278	6.180	6.380	472.370	500.000	-5.5
Aroclor-1260-2	6.467	6.369	6.569	444.010	500.000	-11.2
Aroclor-1260-3	6.835	6.737	6.937	449.170	500.000	-10.2
Aroclor-1260-4	7.096	6.998	7.198	471.630	500.000	-5.7
Aroclor-1260-5	7.337	7.239	7.439	465.660	500.000	-6.9
Decachlorobiphenyl	8.742	8.645	8.845	54.230	50.000	8.5
Tetrachloro-m-xylene	3.690	3.592	3.792	47.340	50.000	-5.3

CALIBRATION VERIFICATION SUMMARY

Contract: ROYF02

Lab Code: CHEM Case No.: Q1664 SAS No.: Q1664 SDG NO.: Q1664

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 03/18/2025 03/18/2025

Client Sample No.: CCAL04 Date Analyzed: 04/01/2025

Lab Sample No.: AR1660CCC500 Data File : PO110147.D Time Analyzed: 13:47

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.768	4.672	4.872	454.540	500.000	-9.1
Aroclor-1016-2	4.787	4.691	4.891	454.420	500.000	-9.1
Aroclor-1016-3	4.963	4.867	5.067	455.730	500.000	-8.9
Aroclor-1016-4	5.005	4.909	5.109	400.280	500.000	-19.9
Aroclor-1016-5	5.218	5.122	5.322	495.090	500.000	-1.0
Aroclor-1260-1	6.250	6.155	6.355	457.240	500.000	-8.6
Aroclor-1260-2	6.438	6.342	6.542	445.890	500.000	-10.8
Aroclor-1260-3	6.590	6.495	6.695	426.900	500.000	-14.6
Aroclor-1260-4	7.062	6.967	7.167	431.110	500.000	-13.8
Aroclor-1260-5	7.303	7.207	7.407	427.040	500.000	-14.6
Decachlorobiphenyl	8.691	8.596	8.796	45.720	50.000	-8.6
Tetrachloro-m-xylene	3.687	3.590	3.790	45.960	50.000	-8.1