

CALIBRATION VERIFICATION SUMMARY

Contract: ROYF02

Lab Code: CHEM Case No.: Q1664 SAS No.: Q1664 SDG NO.: Q1664

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 03/27/2025 03/27/2025

Client Sample No.: CCAL05 Date Analyzed: 03/31/2025

Lab Sample No.: AR1660CCC500 Data File : PP071018.D Time Analyzed: 12:03

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.676	5.563	5.763	447.090	500.000	-10.6
Aroclor-1016-2	5.696	5.585	5.785	421.740	500.000	-15.7
Aroclor-1016-3	5.760	5.648	5.848	433.330	500.000	-13.3
Aroclor-1016-4	5.857	5.745	5.945	477.290	500.000	-4.5
Aroclor-1016-5	6.151	6.038	6.238	492.100	500.000	-1.6
Aroclor-1260-1	7.271	7.156	7.356	477.250	500.000	-4.6
Aroclor-1260-2	7.525	7.410	7.610	494.750	500.000	-1.1
Aroclor-1260-3	7.884	7.768	7.968	468.080	500.000	-6.4
Aroclor-1260-4	8.108	7.992	8.192	441.170	500.000	-11.8
Aroclor-1260-5	8.429	8.311	8.511	505.450	500.000	1.1
Decachlorobiphenyl	10.251	10.125	10.325	51.750	50.000	3.5
Tetrachloro-m-xylene	4.521	4.412	4.612	49.370	50.000	-1.3

CALIBRATION VERIFICATION SUMMARY

Contract: ROYF02

Lab Code: CHEM Case No.: Q1664 SAS No.: Q1664 SDG NO.: Q1664

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 03/27/2025 03/27/2025

Client Sample No.: CCAL05 Date Analyzed: 03/31/2025

Lab Sample No.: AR1660CCC500 Data File : PP071018.D Time Analyzed: 12:03

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	4.905	4.803	5.003	500.540	500.000	0.1
Aroclor-1016-2	4.923	4.822	5.022	485.520	500.000	-2.9
Aroclor-1016-3	5.101	4.999	5.199	496.340	500.000	-0.7
Aroclor-1016-4	5.142	5.041	5.241	513.350	500.000	2.7
Aroclor-1016-5	5.358	5.255	5.455	505.030	500.000	1.0
Aroclor-1260-1	6.394	6.292	6.492	515.610	500.000	3.1
Aroclor-1260-2	6.583	6.480	6.680	509.330	500.000	1.9
Aroclor-1260-3	6.736	6.634	6.834	501.920	500.000	0.4
Aroclor-1260-4	7.208	7.105	7.305	479.580	500.000	-4.1
Aroclor-1260-5	7.449	7.346	7.546	452.290	500.000	-9.5
Decachlorobiphenyl	8.863	8.759	8.959	45.520	50.000	-9.0
Tetrachloro-m-xylene	3.817	3.716	3.916	49.450	50.000	-1.1